

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4103 SAS No.: P4103 SDG NO.: P4103

Continuing Calib Date: 09/19/2024 Initial Calibration Date(s): 09/09/2024 09/09/2024

Continuing Calib Time: 09:17 Initial Calibration Time(s): 15:31 23:20

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.62	5.62	5.52	5.72	0.00
Aroclor-1016-2	(2)	5.64	5.64	5.54	5.74	0.00
Aroclor-1016-3	(3)	5.70	5.70	5.60	5.80	0.00
Aroclor-1016-4	(4)	5.80	5.80	5.70	5.90	0.00
Aroclor-1016-5	(5)	6.09	6.10	6.00	6.20	0.01
Aroclor-1260-1	(1)	7.22	7.22	7.12	7.32	0.00
Aroclor-1260-2	(2)	7.47	7.48	7.38	7.58	0.01
Aroclor-1260-3	(3)	7.83	7.83	7.73	7.93	0.00
Aroclor-1260-4	(4)	8.06	8.06	7.96	8.16	0.00
Aroclor-1260-5	(5)	8.38	8.38	8.28	8.48	0.00
Tetrachloro-m-xylene		4.45	4.45	4.35	4.55	0.00
Decachlorobiphenyl		10.20	10.20	10.10	10.30	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4103 SAS No.: P4103 SDG NO.: P4103

Continuing Calib Date: 09/19/2024 Initial Calibration Date(s): 09/09/2024 09/09/2024

Continuing Calib Time: 09:17 Initial Calibration Time(s): 15:31 23:20

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.80	4.81	4.71	4.91	0.01
Aroclor-1016-2	(2)	4.82	4.83	4.73	4.93	0.01
Aroclor-1016-3	(3)	5.00	5.00	4.90	5.10	0.00
Aroclor-1016-4	(4)	5.04	5.04	4.94	5.14	0.00
Aroclor-1016-5	(5)	5.25	5.26	5.16	5.36	0.01
Aroclor-1260-1	(1)	6.29	6.29	6.19	6.39	0.01
Aroclor-1260-2	(2)	6.47	6.48	6.38	6.58	0.01
Aroclor-1260-3	(3)	6.63	6.63	6.53	6.73	0.00
Aroclor-1260-4	(4)	7.10	7.11	7.01	7.21	0.01
Aroclor-1260-5	(5)	7.34	7.35	7.25	7.45	0.01
Tetrachloro-m-xylene		3.71	3.72	3.62	3.82	0.01
Decachlorobiphenyl		8.74	8.75	8.65	8.85	0.01