

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4103 SAS No.: P4103 SDG NO.: P4103

Continuing Calib Date: 09/19/2024 Initial Calibration Date(s): 09/09/2024 09/09/2024

Continuing Calib Time: 09:53 Initial Calibration Time(s): 15:31 23:20

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	5.62	5.62	5.52	5.72	0.01
Aroclor-1248-2	(2)	5.89	5.89	5.79	5.99	0.00
Aroclor-1248-3	(3)	6.09	6.09	5.99	6.19	0.00
Aroclor-1248-4	(4)	6.50	6.50	6.40	6.60	0.01
Aroclor-1248-5	(5)	6.53	6.53	6.43	6.63	0.00
Tetrachloro-m-xylene		4.45	4.45	4.35	4.55	0.00
Decachlorobiphenyl		10.20	10.20	10.10	10.30	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4103 SAS No.: P4103 SDG NO.: P4103

Continuing Calib Date: 09/19/2024 Initial Calibration Date(s): 09/09/2024 09/09/2024

Continuing Calib Time: 09:53 Initial Calibration Time(s): 15:31 23:20

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.80	4.80	4.70	4.90	0.00
Aroclor-1248-2 (2)	5.04	5.04	4.94	5.14	0.00
Aroclor-1248-3 (3)	5.08	5.08	4.98	5.18	0.00
Aroclor-1248-4 (4)	5.25	5.26	5.16	5.36	0.01
Aroclor-1248-5 (5)	5.65	5.65	5.55	5.75	0.01
Tetrachloro-m-xylene	3.71	3.72	3.62	3.82	0.01
Decachlorobiphenyl	8.74	8.74	8.64	8.84	0.00