

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4103 SAS No.: P4103 SDG NO.: P4103

Continuing Calib Date: 09/19/2024 Initial Calibration Date(s): 09/09/2024 09/09/2024

Continuing Calib Time: 20:58 Initial Calibration Time(s): 15:31 23:20

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	6.47	6.47	6.37	6.57	0.00
Aroclor-1254-2	(2)	6.69	6.69	6.59	6.79	0.00
Aroclor-1254-3	(3)	7.05	7.05	6.95	7.15	0.00
Aroclor-1254-4	(4)	7.34	7.34	7.24	7.44	0.00
Aroclor-1254-5	(5)	7.76	7.76	7.66	7.86	0.00
Tetrachloro-m-xylene		4.45	4.45	4.35	4.55	0.00
Decachlorobiphenyl		10.19	10.19	10.09	10.29	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4103 SAS No.: P4103 SDG NO.: P4103

Continuing Calib Date: 09/19/2024 Initial Calibration Date(s): 09/09/2024 09/09/2024

Continuing Calib Time: 20:58 Initial Calibration Time(s): 15:31 23:20

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.60	5.61	5.51	5.71	0.01
Aroclor-1254-2 (2)	5.75	5.76	5.66	5.86	0.01
Aroclor-1254-3 (3)	6.16	6.16	6.06	6.26	0.00
Aroclor-1254-4 (4)	6.38	6.39	6.29	6.49	0.01
Aroclor-1254-5 (5)	6.80	6.81	6.71	6.91	0.01
Tetrachloro-m-xylene	3.71	3.72	3.62	3.82	0.01
Decachlorobiphenyl	8.74	8.74	8.64	8.84	0.00