

### CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4103 SAS No.: P4103 SDG NO.: P4103

Continuing Calib Date: 09/19/2024 Initial Calibration Date(s): 09/09/2024 09/09/2024

Continuing Calib Time: 23:48 Initial Calibration Time(s): 15:31 23:20

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1254-1       | (1) | 6.47       | 6.47      | 6.37      | 6.57  | 0.00       |
| Aroclor-1254-2       | (2) | 6.69       | 6.69      | 6.59      | 6.79  | 0.00       |
| Aroclor-1254-3       | (3) | 7.05       | 7.05      | 6.95      | 7.15  | 0.00       |
| Aroclor-1254-4       | (4) | 7.34       | 7.34      | 7.24      | 7.44  | 0.00       |
| Aroclor-1254-5       | (5) | 7.76       | 7.76      | 7.66      | 7.86  | 0.00       |
| Tetrachloro-m-xylene |     | 4.45       | 4.45      | 4.35      | 4.55  | 0.00       |
| Decachlorobiphenyl   |     | 10.20      | 10.19     | 10.09     | 10.29 | -0.01      |

### CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4103 SAS No.: P4103 SDG NO.: P4103

Continuing Calib Date: 09/19/2024 Initial Calibration Date(s): 09/09/2024 09/09/2024

Continuing Calib Time: 23:48 Initial Calibration Time(s): 15:31 23:20

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.60       | 5.61      | 5.51      | 5.71 | 0.01       |
| Aroclor-1254-2 (2)   | 5.75       | 5.76      | 5.66      | 5.86 | 0.01       |
| Aroclor-1254-3 (3)   | 6.16       | 6.16      | 6.06      | 6.26 | 0.00       |
| Aroclor-1254-4 (4)   | 6.38       | 6.39      | 6.29      | 6.49 | 0.01       |
| Aroclor-1254-5 (5)   | 6.80       | 6.81      | 6.71      | 6.91 | 0.01       |
| Tetrachloro-m-xylene | 3.71       | 3.72      | 3.62      | 3.82 | 0.01       |
| Decachlorobiphenyl   | 8.74       | 8.74      | 8.64      | 8.84 | 0.00       |