

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4103 SAS No.: P4103 SDG NO.: P4103

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/09/2024 09/09/2024

Client Sample No.: CCAL01 Date Analyzed: 09/19/2024

Lab Sample No.: AR1660CCC500 Data File : PO106689.D Time Analyzed: 14:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.616	5.517	5.717	573.610	500.000	14.7
Aroclor-1016-2	5.639	5.540	5.740	577.330	500.000	15.5
Aroclor-1016-3	5.701	5.602	5.802	578.600	500.000	15.7
Aroclor-1016-4	5.799	5.701	5.901	581.470	500.000	16.3
Aroclor-1016-5	6.093	5.995	6.195	580.420	500.000	16.1
Aroclor-1260-1	7.218	7.119	7.319	573.860	500.000	14.8
Aroclor-1260-2	7.474	7.375	7.575	555.450	500.000	11.1
Aroclor-1260-3	7.833	7.734	7.934	527.570	500.000	5.5
Aroclor-1260-4	8.059	7.959	8.159	527.520	500.000	5.5
Aroclor-1260-5	8.378	8.279	8.479	522.640	500.000	4.5
Decachlorobiphenyl	10.200	10.097	10.297	51.670	50.000	3.3
Tetrachloro-m-xylene	4.452	4.354	4.554	58.870	50.000	17.7

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4103 SAS No.: P4103 SDG NO.: P4103

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/09/2024 09/09/2024

Client Sample No.: CCAL01 Date Analyzed: 09/19/2024

Lab Sample No.: AR1660CCC500 Data File : PO106689.D Time Analyzed: 14:10

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.800	4.706	4.906	520.620	500.000	4.1
Aroclor-1016-2	4.819	4.726	4.926	517.160	500.000	3.4
Aroclor-1016-3	4.996	4.903	5.103	519.220	500.000	3.8
Aroclor-1016-4	5.036	4.943	5.143	512.890	500.000	2.6
Aroclor-1016-5	5.252	5.158	5.358	555.520	500.000	11.1
Aroclor-1260-1	6.286	6.193	6.393	534.600	500.000	6.9
Aroclor-1260-2	6.472	6.380	6.580	536.260	500.000	7.3
Aroclor-1260-3	6.627	6.534	6.734	521.060	500.000	4.2
Aroclor-1260-4	7.100	7.007	7.207	547.090	500.000	9.4
Aroclor-1260-5	7.339	7.245	7.445	548.730	500.000	9.7
Decachlorobiphenyl	8.738	8.646	8.846	54.410	50.000	8.8
Tetrachloro-m-xylene	3.713	3.617	3.817	53.200	50.000	6.4