

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4103 SAS No.: P4103 SDG NO.: P4103

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/09/2024 09/09/2024

Client Sample No.: CCAL02 Date Analyzed: 09/19/2024

Lab Sample No.: AR1660CCC500 Data File : PO106704.D Time Analyzed: 20:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.614	5.517	5.717	543.470	500.000	8.7
Aroclor-1016-2	5.637	5.540	5.740	544.110	500.000	8.8
Aroclor-1016-3	5.699	5.602	5.802	544.460	500.000	8.9
Aroclor-1016-4	5.797	5.701	5.901	545.010	500.000	9.0
Aroclor-1016-5	6.091	5.995	6.195	536.340	500.000	7.3
Aroclor-1260-1	7.216	7.119	7.319	541.900	500.000	8.4
Aroclor-1260-2	7.472	7.375	7.575	521.450	500.000	4.3
Aroclor-1260-3	7.831	7.734	7.934	478.320	500.000	-4.3
Aroclor-1260-4	8.057	7.959	8.159	469.090	500.000	-6.2
Aroclor-1260-5	8.376	8.279	8.479	488.310	500.000	-2.3
Decachlorobiphenyl	10.195	10.097	10.297	50.390	50.000	0.8
Tetrachloro-m-xylene	4.450	4.354	4.554	56.460	50.000	12.9

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4103 SAS No.: P4103 SDG NO.: P4103

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/09/2024 09/09/2024

Client Sample No.: CCAL02 Date Analyzed: 09/19/2024

Lab Sample No.: AR1660CCC500 Data File : PO106704.D Time Analyzed: 20:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.800	4.706	4.906	514.400	500.000	2.9
Aroclor-1016-2	4.819	4.726	4.926	507.980	500.000	1.6
Aroclor-1016-3	4.996	4.903	5.103	508.640	500.000	1.7
Aroclor-1016-4	5.036	4.943	5.143	501.030	500.000	0.2
Aroclor-1016-5	5.251	5.158	5.358	553.270	500.000	10.7
Aroclor-1260-1	6.286	6.193	6.393	512.960	500.000	2.6
Aroclor-1260-2	6.473	6.380	6.580	516.460	500.000	3.3
Aroclor-1260-3	6.627	6.534	6.734	496.550	500.000	-0.7
Aroclor-1260-4	7.099	7.007	7.207	512.250	500.000	2.5
Aroclor-1260-5	7.339	7.245	7.445	513.530	500.000	2.7
Decachlorobiphenyl	8.737	8.646	8.846	54.190	50.000	8.4
Tetrachloro-m-xylene	3.713	3.617	3.817	51.840	50.000	3.7