

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4103 SAS No.: P4103 SDG NO.: P4103

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/09/2024 09/09/2024

Client Sample No.: CCAL03 Date Analyzed: 09/19/2024

Lab Sample No.: AR1660CCC500 Data File : PO106711.D Time Analyzed: 22:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.616	5.517	5.717	559.830	500.000	12.0
Aroclor-1016-2	5.639	5.540	5.740	560.080	500.000	12.0
Aroclor-1016-3	5.701	5.602	5.802	557.720	500.000	11.5
Aroclor-1016-4	5.799	5.701	5.901	561.280	500.000	12.3
Aroclor-1016-5	6.093	5.995	6.195	554.230	500.000	10.8
Aroclor-1260-1	7.218	7.119	7.319	578.000	500.000	15.6
Aroclor-1260-2	7.473	7.375	7.575	556.260	500.000	11.3
Aroclor-1260-3	7.833	7.734	7.934	507.390	500.000	1.5
Aroclor-1260-4	8.058	7.959	8.159	491.400	500.000	-1.7
Aroclor-1260-5	8.378	8.279	8.479	503.860	500.000	0.8
Decachlorobiphenyl	10.198	10.097	10.297	51.420	50.000	2.8
Tetrachloro-m-xylene	4.452	4.354	4.554	56.700	50.000	13.4

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4103 SAS No.: P4103 SDG NO.: P4103

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/09/2024 09/09/2024

Client Sample No.: CCAL03 Date Analyzed: 09/19/2024

Lab Sample No.: AR1660CCC500 Data File : PO106711.D Time Analyzed: 22:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.801	4.706	4.906	508.700	500.000	1.7
Aroclor-1016-2	4.820	4.726	4.926	508.390	500.000	1.7
Aroclor-1016-3	4.997	4.903	5.103	504.670	500.000	0.9
Aroclor-1016-4	5.037	4.943	5.143	495.190	500.000	-1.0
Aroclor-1016-5	5.252	5.158	5.358	502.020	500.000	0.4
Aroclor-1260-1	6.287	6.193	6.393	523.240	500.000	4.6
Aroclor-1260-2	6.474	6.380	6.580	527.130	500.000	5.4
Aroclor-1260-3	6.628	6.534	6.734	508.980	500.000	1.8
Aroclor-1260-4	7.101	7.007	7.207	522.520	500.000	4.5
Aroclor-1260-5	7.340	7.245	7.445	527.390	500.000	5.5
Decachlorobiphenyl	8.739	8.646	8.846	55.480	50.000	11.0
Tetrachloro-m-xylene	3.713	3.617	3.817	51.860	50.000	3.7