

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4103

Client: ENTACT

Analytical Method: 8082A

DataFile : PO106701.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	WC-14-5-7.5BMSD											
P4103-07MSD	AR1016	184.6	0	140	ug/kg	76		1		40 (55)	140 (146)	30 (20)
	AR1260	184.6	0	177	ug/kg	96		3		40 (45)	140 (144)	30 (20)