

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4103 SAS No.: P4103 SDG NO.: P4103

Continuing Calib Date: 09/23/2024 Initial Calibration Date(s): 09/23/2024 09/23/2024

Continuing Calib Time: 16:40 Initial Calibration Time(s): 11:32 12:26

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.06	9.06	8.96	9.16	0.00
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.92	4.92	4.82	5.02	0.00
Heptachlor epoxide	5.69	5.69	5.59	5.79	0.00
Endrin	6.58	6.58	6.48	6.68	0.00
Methoxychlor	7.50	7.50	7.40	7.60	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4103 SAS No.: P4103 SDG NO.: P4103

Continuing Calib Date: 09/23/2024 Initial Calibration Date(s): 09/23/2024 09/23/2024

Continuing Calib Time: 16:40 Initial Calibration Time(s): 11:32 12:26

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.92	7.92	7.82	8.02	0.00
Tetrachloro-m-xylene	2.78	2.78	2.68	2.88	0.00
gamma-BHC (Lindane)	3.62	3.62	3.52	3.72	0.00
Heptachlor	3.96	3.96	3.86	4.06	0.01
Heptachlor epoxide	4.74	4.74	4.64	4.84	0.00
Endrin	5.65	5.65	5.55	5.75	0.00
Methoxychlor	6.62	6.62	6.52	6.72	0.00