

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4103 SAS No.: P4103 SDG NO.: P4103

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/23/2024 09/23/2024

Client Sample No.: CCAL01 Date Analyzed: 09/23/2024

Lab Sample No.: PSTDCCC050 Data File : PL091976.D Time Analyzed: 16:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.060	8.960	9.160	49.110	50.000	-1.8
Endrin	6.579	6.479	6.679	48.110	50.000	-3.8
gamma-BHC (Lindane)	4.332	4.232	4.432	48.810	50.000	-2.4
Heptachlor	4.921	4.820	5.020	48.770	50.000	-2.5
Heptachlor epoxide	5.689	5.589	5.789	48.550	50.000	-2.9
Methoxychlor	7.504	7.404	7.604	49.420	50.000	-1.2
Tetrachloro-m-xylene	3.543	3.444	3.644	49.440	50.000	-1.1

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4103 SAS No.: P4103 SDG NO.: P4103

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/23/2024 09/23/2024

Client Sample No.: CCAL01 Date Analyzed: 09/23/2024

Lab Sample No.: PSTDCCC050 Data File : PL091976.D Time Analyzed: 16:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.924	7.823	8.023	49.930	50.000	-0.1
Endrin	5.648	5.548	5.748	50.800	50.000	1.6
gamma-BHC (Lindane)	3.616	3.516	3.716	52.480	50.000	5.0
Heptachlor	3.955	3.855	4.055	51.560	50.000	3.1
Heptachlor epoxide	4.738	4.638	4.838	51.340	50.000	2.7
Methoxychlor	6.622	6.521	6.721	50.730	50.000	1.5
Tetrachloro-m-xylene	2.782	2.682	2.882	51.650	50.000	3.3