

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4103 SAS No.: P4103 SDG NO.: P4103

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/23/2024 09/23/2024

Client Sample No.: CCAL02 Date Analyzed: 09/23/2024

Lab Sample No.: PSTDCCC050 Data File : PL091989.D Time Analyzed: 19:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.060	8.960	9.160	48.570	50.000	-2.9
Endrin	6.578	6.479	6.679	48.570	50.000	-2.9
gamma-BHC (Lindane)	4.332	4.232	4.432	47.920	50.000	-4.2
Heptachlor	4.921	4.820	5.020	49.250	50.000	-1.5
Heptachlor epoxide	5.689	5.589	5.789	47.570	50.000	-4.9
Methoxychlor	7.504	7.404	7.604	48.520	50.000	-3.0
Tetrachloro-m-xylene	3.544	3.444	3.644	49.060	50.000	-1.9

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4103 SAS No.: P4103 SDG NO.: P4103

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/23/2024 09/23/2024

Client Sample No.: CCAL02 Date Analyzed: 09/23/2024

Lab Sample No.: PSTDCCC050 Data File : PL091989.D Time Analyzed: 19:35

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.924	7.823	8.023	50.100	50.000	0.2
Endrin	5.648	5.548	5.748	52.220	50.000	4.4
gamma-BHC (Lindane)	3.616	3.516	3.716	51.120	50.000	2.2
Heptachlor	3.955	3.855	4.055	52.290	50.000	4.6
Heptachlor epoxide	4.738	4.638	4.838	50.440	50.000	0.9
Methoxychlor	6.621	6.521	6.721	50.650	50.000	1.3
Tetrachloro-m-xylene	2.782	2.682	2.882	50.940	50.000	1.9