

Surrogate Summary

SDG No.: **P4103**

Client: **ENTACT**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL091953.D	PIBLK-PL091953.D	Decachlorobiphenyl	1	20	19.6	98		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	19.0	95		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	18.7	94		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	18.3	92		30 (77)	150 (126)
I.BLK-PL091974.D	PIBLK-PL091974.D	Decachlorobiphenyl	1	20	19.8	99		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	19.3	96		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	19.6	98		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	18.9	94		30 (77)	150 (126)
PB163572BL	PB163572BL	Decachlorobiphenyl	1	20	19.4	97		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	18.6	93		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	19.1	96		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	18.2	91		30 (77)	150 (126)
PB163572BS	PB163572BS	Decachlorobiphenyl	1	20	18.1	90		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	18.1	90		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	17.9	89		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	16.9	85		30 (77)	150 (126)
P4103-02	WC-22A	Decachlorobiphenyl	1	20	16.9	85		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	13.8	69		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	17.1	86		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	20.4	102		30 (77)	150 (126)
P4103-02MS	WC-22AMS	Decachlorobiphenyl	1	20	18.3	91		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	13.0	65		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	18.6	93		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	21.8	109		30 (77)	150 (126)
P4103-02MSD	WC-22AMSD	Decachlorobiphenyl	1	20	18.0	90		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	12.1	61		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	18.2	91		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	20.8	104		30 (77)	150 (126)
P4103-05	WC-14-5-7.5A	Decachlorobiphenyl	1	20	13.8	69		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	17.3	87		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	14.2	71		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	19.6	98		30 (77)	150 (126)
P4103-08	WC-14-5-7.5B	Decachlorobiphenyl	1	20	14.0	70		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	17.2	86		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	14.3	72		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	18.1	90		30 (77)	150 (126)
P4103-11	WC-14-7.5-10A	Decachlorobiphenyl	1	20	14.4	72		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	16.2	81		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	14.5	72		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	21.0	105		30 (77)	150 (126)
P4103-14	WC-14-7.5-10B	Decachlorobiphenyl	1	20	16.3	82		30 (43)	150 (140)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: **P4103**

Client: **ENTACT**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
P4103-14	WC-14-7.5-10B	Tetrachloro-m-xylene	1	20	15.7	78		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	16.6	83		30 (43)	150 (140)
P4103-17	WC-23	Tetrachloro-m-xylene	2	20	18.3	91		30 (77)	150 (126)
		Decachlorobiphenyl	1	20	20.3	101		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	17.6	88		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	20.4	102		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	20.4	102		30 (77)	150 (126)
PB163511TB	PB163511TB	Decachlorobiphenyl	1	20	19.4	97		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	18.8	94		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	19.2	96		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	18.4	92		30 (77)	150 (126)
I.BLK-PL091988.D	PIBLK-PL091988.D	Decachlorobiphenyl	1	20	19.7	99		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	18.9	95		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	19.4	97		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	18.7	94		30 (77)	150 (126)