

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091924\
 Data File : BP021956.D
 Acq On : 19 Sep 2024 18:25
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP091924

Quant Time: Sep 20 01:25:53 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 20 01:22:00 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	116	0.00
2	1,4-Dioxane	40.000	37.493	6.3	111	0.00
3	Pyridine	40.000	37.681	5.8	110	0.00
4	n-Nitrosodimethylamine	40.000	37.104	7.2	110	0.00
5 S	2-Fluorophenol	80.000	77.466	3.2	113	0.00
6	Aniline	40.000	39.196	2.0	114	0.00
7 S	Phenol-d6	80.000	78.189	2.3	115	0.00
8	2-Chlorophenol	40.000	39.137	2.2	113	0.00
9	Benzaldehyde	40.000	40.677	-1.7	116	0.00
10 C	Phenol	40.000	38.803	3.0	114	0.00
11	bis(2-Chloroethyl)ether	40.000	39.455	1.4	117	0.00
12	1,3-Dichlorobenzene	40.000	39.303	1.7	115	0.00
13 C	1,4-Dichlorobenzene	40.000	39.031	2.4	114	0.00
14	1,2-Dichlorobenzene	40.000	38.285	4.3	111	0.00
15	Benzyl Alcohol	40.000	40.269	-0.7	116	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.025	-0.1	117	0.00
17	2-Methylphenol	40.000	40.367	-0.9	114	0.00
18	Hexachloroethane	40.000	39.011	2.5	113	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.079	-2.7	116	0.00
20	3+4-Methylphenols	40.000	39.936	0.2	115	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	111	0.00
22	Acetophenone	40.000	39.380	1.5	114	0.00
23 S	Nitrobenzene-d5	80.000	80.399	-0.5	114	0.00
24	Nitrobenzene	40.000	40.119	-0.3	113	0.00
25	Isophorone	40.000	41.517	-3.8	116	0.00
26 C	2-Nitrophenol	40.000	41.304	-3.3	115	0.00
27	2,4-Dimethylphenol	40.000	41.257	-3.1	116	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.255	-0.6	114	0.00
29 C	2,4-Dichlorophenol	40.000	40.661	-1.7	111	0.00
30	1,2,4-Trichlorobenzene	40.000	39.410	1.5	112	0.00
31	Naphthalene	40.000	39.645	0.9	113	0.00
32	Benzoic acid	40.000	41.608	-4.0	112	0.00
33	4-Chloroaniline	40.000	40.065	-0.2	111	0.00
34 C	Hexachlorobutadiene	40.000	39.651	0.9	114	0.01
35	Caprolactam	40.000	40.666	-1.7	112	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.590	-1.5	113	0.00
37	2-Methylnaphthalene	40.000	40.104	-0.3	114	0.00
38	1-Methylnaphthalene	40.000	39.551	1.1	113	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	115	0.01
40	1,2,4,5-Tetrachlorobenzene	40.000	38.509	3.7	113	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.303	1.7	116	0.00
42 S	2,4,6-Tribromophenol	80.000	81.916	-2.4	121	-0.01
43 C	2,4,6-Trichlorophenol	40.000	40.220	-0.5	115	0.00
44	2,4,5-Trichlorophenol	40.000	40.069	-0.2	116	0.00
45 S	2-Fluorobiphenyl	80.000	77.828	2.7	116	0.00
46	1,1'-Biphenyl	40.000	38.916	2.7	116	0.00
47	2-Chloronaphthalene	40.000	39.460	1.3	116	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091924\
 Data File : BP021956.D
 Acq On : 19 Sep 2024 18:25
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP091924

Quant Time: Sep 20 01:25:53 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 20 01:22:00 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.476	-3.7	114	0.00
49	Acenaphthylene	40.000	39.338	1.7	113	0.00
50	Dimethylphthalate	40.000	39.851	0.4	116	0.00
51	2,6-Dinitrotoluene	40.000	40.539	-1.3	116	0.01
52 C	Acenaphthene	40.000	39.066	2.3	115	0.00
53	3-Nitroaniline	40.000	40.571	-1.4	112	0.00
54 P	2,4-Dinitrophenol	40.000	41.913	-4.8	118	0.00
55	Dibenzofuran	40.000	39.346	1.6	116	0.00
56 P	4-Nitrophenol	40.000	41.208	-3.0	111	0.01
57	2,4-Dinitrotoluene	40.000	41.226	-3.1	116	0.00
58	Fluorene	40.000	39.205	2.0	115	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.794	0.5	116	0.00
60	Diethylphthalate	40.000	40.374	-0.9	119	0.00
61	4-Chlorophenyl-phenylether	40.000	39.472	1.3	118	0.00
62	4-Nitroaniline	40.000	40.580	-1.4	112	0.00
63	Azobenzene	40.000	39.347	1.6	115	0.01
64 I	Phenanthrene-d10	20.000	20.000	0.0	115	-0.01
65	4,6-Dinitro-2-methylphenol	40.000	42.104	-5.3	118	-0.02
66 c	n-Nitrosodiphenylamine	40.000	40.122	-0.3	115	0.01
67	4-Bromophenyl-phenylether	40.000	40.562	-1.4	120	0.00
68	Hexachlorobenzene	40.000	40.247	-0.6	119	0.00
69	Atrazine	40.000	35.859	10.4	115	0.00
70 C	Pentachlorophenol	40.000	40.450	-1.1	115	0.00
71	Phenanthrene	40.000	39.089	2.3	115	-0.01
72	Anthracene	40.000	39.418	1.5	114	0.00
73	Carbazole	40.000	39.988	0.0	113	0.00
74	Di-n-butylphthalate	40.000	42.142	-5.4	125	0.00
75 C	Fluoranthene	40.000	40.668	-1.7	118	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	114	0.00
77	Benzidine	40.000	57.157	-42.9#	99	0.00
78	Pyrene	40.000	40.184	-0.5	114	0.00
79 S	Terphenyl-d14	80.000	81.703	-2.1	117	0.00
80	Butylbenzylphthalate	40.000	43.953	-9.9	125	0.00
81	Benzo(a)anthracene	40.000	40.503	-1.3	118	0.00
82	3,3'-Dichlorobenzidine	40.000	42.500	-6.3	114	0.00
83	Chrysene	40.000	39.565	1.1	115	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.565	-11.4	132	0.00
85 c	Di-n-octyl phthalate	40.000	44.909	-12.3	132	0.00
86 I	Perylene-d12	20.000	20.000	0.0	115	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.370	1.6	114	-0.03
88	Benzo(b)fluoranthene	40.000	39.435	1.4	112	0.00
89	Benzo(k)fluoranthene	40.000	39.807	0.5	117	0.00
90 C	Benzo(a)pyrene	40.000	40.460	-1.2	116	0.00
91	Dibenzo(a,h)anthracene	40.000	39.199	2.0	115	0.00
92	Benzo(g,h,i)perylene	40.000	39.624	0.9	114	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091924\
Data File : BP021956.D
Acq On : 19 Sep 2024 18:25
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP091924

Quant Time: Sep 20 01:25:53 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 20 01:22:00 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0