

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092024\
 Data File : BF139493.D
 Acq On : 20 Sep 2024 10:09
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 20 11:16:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 14 02:58:23 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	138	0.00
2	1,4-Dioxane	40.000	43.111	-7.8	148	-0.02
3	Pyridine	40.000	46.400	-16.0	158	-0.02
4	n-Nitrosodimethylamine	40.000	36.937	7.7	127	-0.02
5 S	2-Fluorophenol	80.000	79.712	0.4	139	0.00
6	Aniline	40.000	40.730	-1.8	143	-0.02
7 S	Phenol-d6	80.000	80.084	-0.1	139	0.00
8	2-Chlorophenol	40.000	40.610	-1.5	143	0.00
9	Benzaldehyde	40.000	34.834	12.9	134	0.00
10 C	Phenol	40.000	41.296	-3.2	145	0.00
11	bis(2-Chloroethyl)ether	40.000	40.387	-1.0	153	0.00
12	1,3-Dichlorobenzene	40.000	39.552	1.1	139	0.00
13 C	1,4-Dichlorobenzene	40.000	40.182	-0.5	141	0.00
14	1,2-Dichlorobenzene	40.000	39.684	0.8	140	0.00
15	Benzyl Alcohol	40.000	38.676	3.3	134	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	44.868	-12.2	159	0.00
17	2-Methylphenol	40.000	41.437	-3.6	146	0.00
18	Hexachloroethane	40.000	38.249	4.4	134	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.259	-0.6	141	0.00
20	3+4-Methylphenols	40.000	39.607	1.0	139	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	148	0.00
22	Acetophenone	40.000	37.910	5.2	136	0.00
23 S	Nitrobenzene-d5	80.000	77.439	3.2	138	0.00
24	Nitrobenzene	40.000	39.014	2.5	139	0.00
25	Isophorone	40.000	40.062	-0.2	145	0.00
26 C	2-Nitrophenol	40.000	42.833	-7.1	148	0.00
27	2,4-Dimethylphenol	40.000	40.804	-2.0	144	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.865	-2.2	149	0.00
29 C	2,4-Dichlorophenol	40.000	40.104	-0.3	143	0.00
30	1,2,4-Trichlorobenzene	40.000	38.526	3.7	139	0.00
31	Naphthalene	40.000	39.615	1.0	143	0.00
32	Benzoic acid	40.000	40.999	-2.5	145	0.00
33	4-Chloroaniline	40.000	40.856	-2.1	143	0.00
34 C	Hexachlorobutadiene	40.000	36.497	8.8	131	0.00
35	Caprolactam	40.000	42.778	-6.9	149	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.626	0.9	140	0.00
37	2-Methylnaphthalene	40.000	39.129	2.2	141	0.00
38	1-Methylnaphthalene	40.000	39.364	1.6	143	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	149	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.909	2.7	139	0.00
41 P	Hexachlorocyclopentadiene	40.000	35.909	10.2	121	0.00
42 S	2,4,6-Tribromophenol	80.000	72.067	9.9	125	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.431	1.4	135	0.00
44	2,4,5-Trichlorophenol	40.000	40.768	-1.9	143	0.00
45 S	2-Fluorobiphenyl	80.000	74.973	6.3	134	0.00
46	1,1'-Biphenyl	40.000	39.108	2.2	139	0.00
47	2-Chloronaphthalene	40.000	39.967	0.1	141	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092024\
 Data File : BF139493.D
 Acq On : 20 Sep 2024 10:09
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 20 11:16:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 14 02:58:23 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.612	-1.5	143	0.00
49	Acenaphthylene	40.000	40.203	-0.5	143	0.00
50	Dimethylphthalate	40.000	38.502	3.7	137	0.00
51	2,6-Dinitrotoluene	40.000	41.341	-3.4	144	0.00
52 C	Acenaphthene	40.000	39.644	0.9	139	0.00
53	3-Nitroaniline	40.000	42.283	-5.7	144	0.00
54 P	2,4-Dinitrophenol	40.000	39.744	0.6	135	0.00
55	Dibenzofuran	40.000	39.547	1.1	140	0.00
56 P	4-Nitrophenol	40.000	39.069	2.3	133	0.00
57	2,4-Dinitrotoluene	40.000	40.802	-2.0	138	0.00
58	Fluorene	40.000	38.221	4.4	133	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.731	3.2	136	0.00
60	Diethylphthalate	40.000	37.413	6.5	131	0.00
61	4-Chlorophenyl-phenylether	40.000	37.365	6.6	132	0.00
62	4-Nitroaniline	40.000	38.069	4.8	130	0.00
63	Azobenzene	40.000	39.229	1.9	137	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	142	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.802	-9.5	141	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.876	0.3	138	0.00
67	4-Bromophenyl-phenylether	40.000	39.164	2.1	134	0.00
68	Hexachlorobenzene	40.000	38.303	4.2	130	0.00
69	Atrazine	40.000	35.465	11.3	145	0.00
70 C	Pentachlorophenol	40.000	38.730	3.2	128	0.00
71	Phenanthrene	40.000	39.694	0.8	136	0.00
72	Anthracene	40.000	40.170	-0.4	136	0.00
73	Carbazole	40.000	40.767	-1.9	138	0.00
74	Di-n-butylphthalate	40.000	38.771	3.1	129	0.00
75 C	Fluoranthene	40.000	40.219	-0.5	136	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	130	0.00
77	Benzydine	40.000	34.715	13.2	85	0.00
78	Pyrene	40.000	47.111	-17.8	137	0.00
79 S	Terphenyl-d14	80.000	86.345	-7.9	127	0.00
80	Butylbenzylphthalate	40.000	42.049	-5.1	128	0.00
81	Benzo(a)anthracene	40.000	39.470	1.3	126	0.00
82	3,3'-Dichlorobenzidine	40.000	36.590	8.5	124	0.00
83	Chrysene	40.000	40.098	-0.2	129	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.547	3.6	122	0.00
85 c	Di-n-octyl phthalate	40.000	35.467	11.3	112	0.00
86 I	Perylene-d12	20.000	20.000	0.0	135	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.144	-7.9	127	0.00
88	Benzo(b)fluoranthene	40.000	39.093	2.3	132	0.00
89	Benzo(k)fluoranthene	40.000	36.857	7.9	118	0.00
90 C	Benzo(a)pyrene	40.000	40.170	-0.4	129	0.00
91	Dibenzo(a,h)anthracene	40.000	42.712	-6.8	129	0.00
92	Benzo(g,h,i)perylene	40.000	45.678	-14.2	135	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092024\
Data File : BF139493.D
Acq On : 20 Sep 2024 10:09
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Sep 20 11:16:14 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 14 02:58:23 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0