

Report of Analysis

Client:	ENTACT		Date Collected:	09/17/24	
Project:	North Point		Date Received:	09/18/24	
Client Sample ID:	WC-22AMSD		SDG No.:	P4103	
Lab Sample ID:	P4103-02MSD		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139507.D	1	09/20/24 11:30	09/20/24 17:52	PB163571

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	360		15.5	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	300		8.40	50.0	ug/L
95-48-7	2-Methylphenol	420		11.3	50.0	ug/L
65794-96-9	3+4-Methylphenols	400		11.5	100	ug/L
67-72-1	Hexachloroethane	280		10.1	50.0	ug/L
98-95-3	Nitrobenzene	390		12.7	50.0	ug/L
87-68-3	Hexachlorobutadiene	340		12.7	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	480		8.90	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	470		10.1	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	440		15.2	50.0	ug/L
118-74-1	Hexachlorobenzene	460		11.4	50.0	ug/L
87-86-5	Pentachlorophenol	1000	E	18.5	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	117		15 (10) - 110 (139)	78%	SPK: 150
13127-88-3	Phenol-d6	114		15 (10) - 110 (134)	76%	SPK: 150
4165-60-0	Nitrobenzene-d5	88.1		30 (49) - 130 (133)	88%	SPK: 100
321-60-8	2-Fluorobiphenyl	92.7		30 (52) - 130 (132)	93%	SPK: 100
118-79-6	2,4,6-Tribromophenol	115		15 (44) - 110 (137)	76%	SPK: 150
1718-51-0	Terphenyl-d14	109		30 (48) - 130 (125)	109%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	71200	6.881			
1146-65-2	Naphthalene-d8	246000	8.163			
15067-26-2	Acenaphthene-d10	113000	9.916			
1517-22-2	Phenanthrene-d10	188000	11.398			
1719-03-5	Chrysene-d12	100000	14.039			
1520-96-3	Perylene-d12	81000	15.51			

Report of Analysis

Client:	ENTACT		Date Collected:	09/17/24	
Project:	North Point		Date Received:	09/18/24	
Client Sample ID:	WC-22AMSD		SDG No.:	P4103	
Lab Sample ID:	P4103-02MSD		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139507.D	1	09/20/24 11:30	09/20/24 17:52	PB163571

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products