

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4103

Client: ENTACT

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	P4103-02MS	Client Sample ID:	WC-22AMS					DataFile:	BF139506.D		
Pyridine	500	0	350	ug/L	70				20 (10)	160 (109)	
1,4-Dichlorobenzene	500	0	310	ug/L	62	*			70 (55)	130 (125)	
2-Methylphenol	500	0	420	ug/L	84				70 (37)	130 (126)	
3+4-Methylphenols	500	0	410	ug/L	82				20 (31)	160 (127)	
Hexachloroethane	500	0	280	ug/L	56				20 (49)	160 (110)	
Nitrobenzene	500	0	370	ug/L	74				70 (62)	130 (112)	
Hexachlorobutadiene	500	0	330	ug/L	66	*			70 (52)	130 (125)	
2,4,6-Trichlorophenol	500	0	470	ug/L	94				70 (78)	130 (112)	
2,4,5-Trichlorophenol	500	0	450	ug/L	90				70 (71)	130 (111)	
2,4-Dinitrotoluene	500	0	450	ug/L	90				70 (50)	130 (142)	
Hexachlorobenzene	500	0	450	ug/L	90				70 (72)	130 (115)	
Pentachlorophenol	1000	93.6	990	ug/L	90				20 (25)	160 (139)	