

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4103

Client: ENTACT

Analytical Method: SW8260D

Datafile : VN084083.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0920WBSD02	Vinyl chloride	20	18.7	ug/L	94	8		70 (65)	130 (117)	20 (20)
	1,1-Dichloroethane	20	20.5	ug/L	103	3		70 (78)	130 (112)	20 (20)
	2-Butanone	100	110	ug/L	110	0		40 (65)	160 (122)	20 (20)
	Carbon Tetrachloride	20	19.7	ug/L	99	2		70 (77)	130 (113)	20 (20)
	Chloroform	20	21.2	ug/L	106	7		70 (79)	130 (113)	20 (20)
	Benzene	20	19.7	ug/L	99	4		70 (82)	130 (109)	20 (20)
	1,2-Dichloroethane	20	20.4	ug/L	102	2		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	19.9	ug/L	100	7		70 (77)	130 (113)	20 (20)
	Toluene	20	19.9	ug/L	100	4		70 (82)	130 (110)	20 (20)
	Tetrachloroethene	20	19.8	ug/L	99	3		70 (67)	130 (123)	20 (20)
	Chlorobenzene	20	19.3	ug/L	97	4		70 (82)	130 (109)	20 (20)
	Ethyl Benzene	20	19.1	ug/L	96	4		70 (83)	130 (109)	20 (20)
	m/p-Xylenes	40	38.7	ug/L	97	5		70 (82)	130 (110)	20 (20)
	o-Xylene	20	19.3	ug/L	97	4		70 (83)	130 (109)	20 (20)