

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: P4103 SAS No.: P4103 SDG NO.: P4103
 Lab File ID: VN084080.D Date Analyzed: 09/21/2024
 Instrument ID: MSVOA_N Time Analyzed: 00:20
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	108072	13.794				
UPPER LIMIT	216144	14.294				
LOWER LIMIT	54036	13.294				
EPA SAMPLE NO.						
WC-22A	94833	13.79				
WC-14-5-7.5A	96079	13.79				
WC-14-5-7.5B	104000	13.79				
WC-14-7.5-10A	106447	13.79				
WC-14-7.5-10B	108316	13.79				
WC-23	107540	13.79				
VN0920WBL03	93481	13.79				
VN0920WBS02	99296	13.79				
VN0920WBSD02	99263	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.