

EPA SAMPLE NO.

PMW-7D(20240927)S

FORM 5A - IN

## MATRIX SPIKE SAMPLE RECOVERY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011

Lab Code: ACE Case No.: ARC Sayrevil MA No. : SDG No.: P4227

Matrix : Water Analytical Method: CVAA

% Solids:

Concentration Units ( $\mu\text{g/L}$ ,  $\text{mg/L}$  or  $\text{mg/kg}$  dry weight):  $\mu\text{g/L}$ 

Analyte	Control Limit %R	Spiked Sample Result (SSR) Q	Sample Result (SR) Q	Spike Added (SA)	%R	Q
Mercury	75 - 125	0.79	0.2 U	1.0	79	