



**284 Sheffield Street  
Mountainside, NJ 07092**

## **SDG NARRATIVE**

**USEPA**

**SDG # P4227**

**CASE # ARC Sayreville PostEx 2024**

**CONTRACT # 68HERH20D0011**

**SOW# SFAM01.1**

**LAB NAME: Alliance Technical Group, LLC**

**LAB CODE: ACE**

**LAB ORDER ID #P4227**

### **A. Number of Samples and Date of Receipt**

20 Water samples were delivered to the laboratory intact on 09/27/2024.

### **B. Parameters**

Test requested for Metals CLP MS = Aluminum, Antimony, Arsenic, Beryllium, Cadmium, Chromium, Iron, Lead, Manganese, Nickel, Thallium and Hg.

### **C. Cooler Temp**

Indicator Bottle: **Presence**/Absence

Cooler: 2.3°C

### **D. Analytical Techniques:**

All analyses were based on CLP Methodology by method SFAM01.1.

### **E. Calculation:**

#### **Calculation for ICP-MS Water Sample:**

$$\text{Concentration or Result } (\mu\text{g/L}) = C \times \frac{V_f}{V_i} \times \text{DF}$$

Where,

C = Instrument value in ppb (The average of all replicate integrations)

V<sub>f</sub> = Final digestion volume (mL)

V<sub>i</sub> = Initial aliquot amount (mL) (Sample amount taken in prep)

DF = Dilution Factor



**284 Sheffield Street  
Mountainside, NJ 07092**

**Example Calculation For Sample PMW-1(20240925) For Lead:**

If C = 0.40 ppb  
Vf = 50 ml  
Vi = 50 ml  
DF = 1

$$\begin{aligned}\text{Concentration or Result } (\mu\text{g/L}) &= 0.40 \times \frac{50}{50} \times 1 \\ &= 0.40 \mu\text{g/L} \\ &= 0.40 \mu\text{g/L (Reported Result with Signification)}\end{aligned}$$

**Calculation for Hg Water Sample:**

$$\text{Concentration or Result } (\mu\text{g/L}) = C \times \text{DF}$$

Where,

C = Instrument response in  $\mu\text{g/L}$  from the calibration curve.  
DF = Dilution Factor

**Example Calculation For Sample PMW-1(20240925):**

If C = 1.3197 ppb  
DF = 1

$$\begin{aligned}\text{Concentration or Result } (\mu\text{g/L}) &= 1.3197 \times 1 \\ &= 1.3197 \mu\text{g/L} \\ &= 1.3 \mu\text{g/L (Reported Result with Signification)}\end{aligned}$$

**F. QA/ QC**

Calibrations met requirements. Interference check met requirements. Blank analyses did not indicate any presence of contamination. Laboratory Control sample was within control limits. Spike sample did meet requirements. Duplicate sample did meet requirements. Serial Dilution did meet requirements.



**284 Sheffield Street  
Mountainside, NJ 07092**

Collision cell is being used to remove potential interferences. The analytes Na, Mg, Al, K, V, Cr, Mn, Fe, Co, Ni, Cu, Zn, As are being analyzed with collision cell and analytes Be, B, Ca, Ti, Se, Sr, Zr, Mo, Ag, Cd, Sn, Sb, Ba, Tl, Pb, U are being analyzed with Non-Collision Cell. Helium gas is used for the Collision Cell analysis.

Internal Standard Association for ICP-MS analysis.

Target Analyte	Associated Internal Standard
Aluminum	45Sc
Antimony	159Tb
Arsenic	89Y
Beryllium	6Li
Cadmium	159Tb
Chromium	45Sc
Iron	45Sc
Lead	209Bi
Manganese	45Sc
Nickel	45Sc
Thallium	209Bi

I certify that the data package is in compliance with the terms and conditions of the contract both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the Laboratory Director or his designee, as verified by the following signature.

Signature\_\_\_\_\_

Name: Nimisha Pandya

Date \_\_\_\_\_

Title: Document Control Officer