

## **CASE NARRATIVE**

**Roman E&G Corp**  
**Project Name: Perth Amboy**  
**Project # N/A**  
**Chemtech Project # P4258**  
**Test Name: Herbicide**

### **A. Number of Samples and Date of Receipt:**

4 Solid samples were received on 10/01/2024.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: Ammonia, Anions Group1, ASTM Ammonia, ASTM COD, ASTM Leach Extraction, ASTM Oil and Grease, ASTM TS, COD, Corrosivity, EPH, GROWS Concrete, GROWS Concrete ASTM, GROWS Soil, GROWS Soil-Waste Class, Herbicide, Hexavalent Chromium, Ignitability, Mercury, Metals Group1, Metals ICP-Group1, Metals ICP-TAL, METALS-TAL, Oil and Grease, Paint Filter, PCB, Pesticide-TCL, pH, Reactive Cyanide, Reactive Sulfide, SVOC-TCL BNA -20, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP Mercury, TCLP Metals + Cu+Ni+Zn, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TCLP-FULL, TS, TVS and VOC-TCLVOA-10. This data package contains results for Herbicide.

### **C. Analytical Techniques:**

The analysis was performed on instrument ECD\_S. The front column is RTX-CLPesticides which is 30 meters, 0.32 mm ID, 0.5 µm df, Catalog # 11139. The rear column is RTX-CLPesticides2 which is 30 meters, 0.32 mm ID, 0.25 µm df, Catalog #: 11324. The analysis of Herbicides was based on method 8151A and extraction was done based on method 3541.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Retention Times were acceptable for all samples.

The MS {P4385-02MS} with File ID: PS027910.D recoveries met the requirements for all compounds except for Dinoseb[0%] due to matrix interference.

The MSD {P4385-02MSD} with File ID: PS027911.D recoveries met the acceptable requirements except for Dinoseb[0%] due to matrix interference.

The RPD for {P4385-02MSD} with File ID: PS027911.D met criteria except for 2,4-DB[47%] due to difference in results of MS and MSD.

The Blank Spike met requirements for all samples .



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

**E. Additional Comments:**

The temperature of the samples at the time of receipt was 22.1°C, Lab notified this issue to the client. See the communication in shipping Document section.

The soil samples results are based on a dry weight basis.

**F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_