

CASE NARRATIVE

Roman E&G Corp

Project Name: Perth Amboy

Project # N/A

Chemtech Project # P4258

Test Name: PCB

A. Number of Samples and Date of Receipt:

4 Solid samples were received on 10/01/2024.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Ammonia, ASTM Ammonia, ASTM COD, ASTM Leach Extraction, ASTM Oil and Grease, ASTM TS, COD, Corrosivity, GROWS Concrete, GROWS Concrete ASTM, GROWS Soil, GROWS Soil-Waste Class, Ignitability, Oil and Grease, Paint Filter, PCB, pH, Reactive Cyanide, Reactive Sulfide, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP Mercury, TCLP Metals + Cu+Ni+Zn, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TCLP-FULL, TS, TVS and VOC-TCLVOA-10. This data package contains results for PCB.

C. Analytical Techniques:

The analyses were performed on instrument GCECD_P. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalogue # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 µm; Catalogue # 7HM-G017-11. The analysis of PCBs was based on method 8082A and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Retention Times were acceptable for all samples.

The MS recoveries met the requirements for all compounds .

The MSD recoveries met the acceptable requirements .

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration File ID PP067441.D met the requirements except for Aroclor-1260(Peak-02),Aroclor-1260(Peak-03) is failing in 1st column but passing in 2nd column therefore no corrective action taken.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

E. Additional Comments:

The temperature of the samples at the time of receipt was 22.1°C, Lab notified this issue to the client. See the communication in shipping Document section.

The soil samples results are based on a dry weight basis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____