

CASE NARRATIVE

Roman E&G Corp

Project Name: Perth Amboy

Project # N/A

Chemtech Project # P4258

Test Name: VOC-TCLVOA-10

A. Number of Samples and Date of Receipt:

4 Solid samples were received on 10/01/2024.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Ammonia, ASTM Ammonia, ASTM COD, ASTM Leach Extraction, ASTM Oil and Grease, ASTM TS, COD, Corrosivity, GROWS Concrete, GROWS Concrete ASTM, GROWS Soil, GROWS Soil-Waste Class, Ignitability, Oil and Grease, Paint Filter, PCB, pH, Reactive Cyanide, Reactive Sulfide, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP Mercury, TCLP Metals + Cu+Ni+Zn, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TCLP-FULL, TS, TVS and VOC-TCLVOA-10. This data package contains results for VOC-TCLVOA-10.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_Y were done using GC column Rxi-624Sil MS, which is 30 meters, 0.25 mm id, 1.4 um df, Restek Cat. #13868. The Trap was supplied by Supelco, VOCARB 3000, ATOMAX XYZ Concentrator. The analysis of VOC-TCLVOA-10 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements except for Chamberlain Ave, Chamberlain Ave RE the failure sample in Internal standard was reanalyzed to confirm the failure as per method and reported.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The %RSD is greater than 15% in the Initial Calibration method (82Y090924S.M) for Dichlorodifluoromethane this compound is passing on Linear Regression.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

The Continuous Calibration File ID VY019756.D met the requirements except for Dichlorodifluoromethane failing marginally low and Bromochloromethane and t-1,3-Dichloropropene failing high but no positive hit in associated sample therefore no corrective action taken.

The Continuous Calibration File ID VY019776.D met the requirements except for Bromochloromethane failing high but no positive hit in associated sample therefore no corrective action taken.

The Tuning criteria met requirements.

E. Additional Comments:

The temperature of the samples at the time of receipt was 22.1°C, Lab notified this issue to the client. See the communication in shipping Document section.

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

Trip Blank was not provided with this set of samples.

The soil samples results are based on a dry weight basis.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____