

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB132914

Review By	jaswal	Review On	10/14/2024 10:26:20 PM
Supervise By	mohan	Supervise On	10/14/2024 10:27:01 PM

STD. NAME	STD REF.#
ICAL Standard	MP82765,MP82766,MP82767,MP82768,MP82769,MP82770
ICV Standard	MP82771
CCV Standard	MP82773
ICSA Standard	
CRI Standard	MP82775
LCS Standard	
Chk Standard	MP82772,MP82774,MP82776,MP82781

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/14/24 11:19		Mohan	OK
2	S0.2	S0.2	CAL2	10/14/24 11:21		Mohan	OK
3	S2.5	S2.5	CAL3	10/14/24 11:23		Mohan	OK
4	S5	S5	CAL4	10/14/24 11:26		Mohan	OK
5	S7.5	S7.5	CAL5	10/14/24 11:28		Mohan	OK
6	S10	S10	CAL6	10/14/24 11:30		Mohan	OK
7	ICV36	ICV36	ICV	10/14/24 11:33		Mohan	OK
8	ICB36	ICB36	ICB	10/14/24 11:38		Mohan	OK
9	CCV13	CCV13	CCV	10/14/24 11:40		Mohan	OK
10	CCB13	CCB13	CCB	10/14/24 11:43		Mohan	OK
11	CRA	CRA	CRDL	10/14/24 11:45		Mohan	OK
12	HighStd	HighStd	HIGH STD	10/14/24 11:47		Mohan	OK
13	ChkStd	ChkStd	SAM	10/14/24 11:49		Mohan	OK
14	PB164114BL	PB164114BL	MB	10/14/24 11:54		Mohan	OK
15	PB164114BS	PB164114BS	LCS	10/14/24 11:57		Mohan	OK
16	P4258-03	Chamberlain Ave	SAM	10/14/24 11:59		Mohan	OK
17	P4364-01	SB-01-1.0-1.5	SAM	10/14/24 12:01		Mohan	OK
18	P4364-02	SB-01-2.5-3.0	SAM	10/14/24 12:03		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB132914

Review By	jaswal	Review On	10/14/2024 10:26:20 PM
Supervise By	mohan	Supervise On	10/14/2024 10:27:01 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82765,MP82766,MP82767,MP82768,MP82769,MP82770		
ICV Standard	MP82771		
CCV Standard	MP82773		
ICSA Standard			
CRI Standard	MP82775		
LCS Standard			
Chk Standard	MP82772,MP82774,MP82776,MP82781		

19	P4364-03	SB-01-4.25-4.83	SAM	10/14/24 12:06		Mohan	OK
20	P4364-04	SB-02-0.167-0.667	SAM	10/14/24 12:08	High	Mohan	Dilution
21	CCV14	CCV14	CCV	10/14/24 12:10		Mohan	OK
22	CCB14	CCB14	CCB	10/14/24 12:13		Mohan	OK
23	P4364-05	SB-02-2.0-2.5	SAM	10/14/24 12:15	High	Mohan	Dilution
24	P4364-06	SB-02-4.5-5.0	SAM	10/14/24 12:17	High	Mohan	Dilution
25	P4364-07	SB-03-0.167-0.667	SAM	10/14/24 12:20	High	Mohan	Dilution
26	P4364-08	SB-03-2.0-2.5	SAM	10/14/24 12:22		Mohan	OK
27	P4364-09	SB-03-4.167-4.667	SAM	10/14/24 12:25		Mohan	OK
28	P4364-10	SB-04-0.167-0.667	SAM	10/14/24 12:27	High	Mohan	Dilution
29	P4364-11	SB-04-2.0-2.5	SAM	10/14/24 12:29		Mohan	OK
30	P4364-12	SB-04-4.5-5.0	SAM	10/14/24 12:32		Mohan	OK
31	P4364-13	SB-05-0.167-0.667	SAM	10/14/24 12:34		Mohan	OK
32	P4364-14	SB-06-0.167-0.667	SAM	10/14/24 12:37		Mohan	OK
33	CCV15	CCV15	CCV	10/14/24 12:39		Mohan	OK
34	CCB15	CCB15	CCB	10/14/24 12:41		Mohan	OK
35	P4364-15	SB-07-0.167-0.667	SAM	10/14/24 12:43		Mohan	OK
36	P4364-16	SB-08-0.167-0.667	SAM	10/14/24 12:46		Mohan	OK
37	P4364-16DUP	SB-08-0.167-0.667DUP	DUP	10/14/24 12:48		Mohan	OK
38	P4364-16MS	SB-08-0.167-0.667MS	MS	10/14/24 12:53		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB132914

Review By	jaswal	Review On	10/14/2024 10:26:20 PM
Supervise By	mohan	Supervise On	10/14/2024 10:27:01 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82765,MP82766,MP82767,MP82768,MP82769,MP82770		
ICV Standard	MP82771		
CCV Standard	MP82773		
ICSA Standard			
CRI Standard	MP82775		
LCS Standard			
Chk Standard	MP82772,MP82774,MP82776,MP82781		

39	P4364-16MSD	SB-08-0.167-0.667MS	MSD	10/14/24 12:55		Mohan	OK
40	P4376-01	OR-02-100924	SAM	10/14/24 12:58		Mohan	OK
41	P4377-01	HD-01-100924	SAM	10/14/24 13:00		Mohan	OK
42	PB164117BL	PB164117BL	MB	10/14/24 13:02		Mohan	OK
43	PB164117BS	PB164117BS	LCS	10/14/24 13:05		Mohan	OK
44	P4375-01	CONCRETE-PADS-1	SAM	10/14/24 13:07		Mohan	OK
45	CCV16	CCV16	CCV	10/14/24 13:09		Mohan	OK
46	CCB16	CCB16	CCB	10/14/24 13:12		Mohan	OK
47	P4375-01DUP	CONCRETE-PADS-1	DUP	10/14/24 13:14		Mohan	OK
48	P4375-01MS	CONCRETE-PADS-1	MS	10/14/24 13:16		Mohan	OK
49	P4375-01MSD	CONCRETE-PADS-1	MSD	10/14/24 13:18		Mohan	OK
50	P4375-03	CONCRETE-PADS-2	SAM	10/14/24 13:21		Mohan	OK
51	P4375-05	CONCRETE-PADS-3	SAM	10/14/24 13:23		Mohan	OK
52	P4375-07	CONCRETE-PADS-4	SAM	10/14/24 13:25		Mohan	OK
53	P4375-09	CONCRETE-PADS-5	SAM	10/14/24 13:28		Mohan	OK
54	P4375-11	CONCRETE-PADS-6	SAM	10/14/24 13:30		Mohan	OK
55	P4375-13	CONCRETE-PADS-7	SAM	10/14/24 13:32		Mohan	OK
56	P4375-15	CONCRETE-PADS-8	SAM	10/14/24 13:34		Mohan	OK
57	CCV17	CCV17	CCV	10/14/24 13:37		Mohan	OK
58	CCB17	CCB17	CCB	10/14/24 13:39		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB132914

Review By	jaswal	Review On	10/14/2024 10:26:20 PM
Supervise By	mohan	Supervise On	10/14/2024 10:27:01 PM

STD. NAME	STD REF.#
ICAL Standard	MP82765,MP82766,MP82767,MP82768,MP82769,MP82770
ICV Standard	MP82771
CCV Standard	MP82773
ICSA Standard	
CRI Standard	MP82775
LCS Standard	
Chk Standard	MP82772,MP82774,MP82776,MP82781

59	P4375-17	CONCRETE-PADS-9	SAM	10/14/24 13:41		Mohan	OK
60	P4375-19	CONCRETE-PADS-10	SAM	10/14/24 13:43		Mohan	OK
61	P4375-21	CONCRETE-PADS-11	SAM	10/14/24 13:46		Mohan	OK
62	P4375-23	FENCE-POST-1	SAM	10/14/24 13:48		Mohan	OK
63	P4375-25	FENCE-POST-2	SAM	10/14/24 13:50		Mohan	OK
64	P4375-27	FENCE-POST-3	SAM	10/14/24 13:52		Mohan	OK
65	PB164118BL	PB164118BL	MB	10/14/24 13:55		Mohan	OK
66	PB164118BS	PB164118BS	LCS	10/14/24 13:57		Mohan	OK
67	P4385-02	SP-1	SAM	10/14/24 13:59		Mohan	OK
68	P4385-04	SP-2	SAM	10/14/24 14:02		Mohan	OK
69	CCV18	CCV18	CCV	10/14/24 14:04		Mohan	OK
70	CCB18	CCB18	CCB	10/14/24 14:06		Mohan	OK
71	P4385-06	SP-3	SAM	10/14/24 14:08		Mohan	OK
72	P4385-08	SP-4	SAM	10/14/24 14:11		Mohan	OK
73	P4385-10	SP-5	SAM	10/14/24 14:13		Mohan	OK
74	P4385-12	SP-6	SAM	10/14/24 14:15		Mohan	OK
75	P4385-14	SP-7	SAM	10/14/24 14:18		Mohan	OK
76	P4385-16	SP-8	SAM	10/14/24 14:20		Mohan	OK
77	P4385-18	SP-9	SAM	10/14/24 14:22		Mohan	OK
78	P4385-20	SP-10	SAM	10/14/24 14:24		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB132914

Review By	jaswal	Review On	10/14/2024 10:26:20 PM
Supervise By	mohan	Supervise On	10/14/2024 10:27:01 PM

STD. NAME	STD REF.#
ICAL Standard	MP82765,MP82766,MP82767,MP82768,MP82769,MP82770
ICV Standard	MP82771
CCV Standard	MP82773
ICSA Standard	
CRI Standard	MP82775
LCS Standard	
Chk Standard	MP82772,MP82774,MP82776,MP82781

79	P4385-20DUP	SP-10DUP	DUP	10/14/24 14:27		Mohan	OK
80	P4385-20MS	SP-10MS	MS	10/14/24 14:29		Mohan	OK
81	CCV19	CCV19	CCV	10/14/24 14:31		Mohan	OK
82	CCB19	CCB19	CCB	10/14/24 14:34		Mohan	OK
83	P4385-20MSD	SP-10MSD	MSD	10/14/24 14:36		Mohan	OK
84	P4393-01	BUST-DEBRIS	SAM	10/14/24 14:38		Mohan	OK
85	P4394-01	WASH-DEBRIS	SAM	10/14/24 14:41		Mohan	OK
86	P4395-01	F05308-SOLID	SAM	10/14/24 14:43		Mohan	OK
87	P4364-16L	SB-08-0.167-0.667L	SD	10/14/24 14:45		Mohan	OK
88	P4364-16A	SB-08-0.167-0.667A	PS	10/14/24 14:47		Mohan	OK
89	P4375-01L	CONCRETE-PADS-1L	SD	10/14/24 14:50		Mohan	OK
90	P4375-01A	CONCRETE-PADS-1A	PS	10/14/24 14:52		Mohan	OK
91	P4385-20L	SP-10L	SD	10/14/24 14:54		Mohan	OK
92	P4385-20A	SP-10A	PS	10/14/24 14:57		Mohan	OK
93	CCV20	CCV20	CCV	10/14/24 14:59		Mohan	OK
94	CCB20	CCB20	CCB	10/14/24 15:01		Mohan	OK
95	P4364-04DL	SB-02-0.167-0.667DL	SAM	10/14/24 15:04	Report 2X	Mohan	Confirms
96	P4364-05DL	SB-02-2.0-2.5DL	SAM	10/14/24 15:06	Report 2X	Mohan	Confirms
97	P4364-06DL	SB-02-4.5-5.0DL	SAM	10/14/24 15:08	Report 5X	Mohan	Confirms
98	P4364-07DL	SB-03-0.167-0.667DL	SAM	10/14/24 15:10	Report 2X	Mohan	Confirms

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QCBatch ID # LB132914

Review By	jaswal	Review On	10/14/2024 10:26:20 PM
Supervise By	mohan	Supervise On	10/14/2024 10:27:01 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82765,MP82766,MP82767,MP82768,MP82769,MP82770		
ICV Standard	MP82771		
CCV Standard	MP82773		
ICSA Standard			
CRI Standard	MP82775		
LCS Standard			
Chk Standard	MP82772,MP82774,MP82776,MP82781		

99	P4364-10DL	SB-04-0.167-0.667DL	SAM	10/14/24 15:13	Report 2X	Mohan	Confirms
100	CCV21	CCV21	CCV	10/14/24 15:15		Mohan	OK
101	CCB21	CCB21	CCB	10/14/24 15:17		Mohan	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132927

Review By	jaswal	Review On	10/21/2024 1:52:41 AM
Supervise By	mohan	Supervise On	10/21/2024 1:54:36 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441,MP82476,MP82477,MP82478,MP82479,MP82484		
ICV Standard	MP82485		
CCV Standard	MP82488		
ICSA Standard	MP82486,MP82487		
CRI Standard	MP82484		
LCS Standard			
Chk Standard	MP82491,MP82492		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/14/24 13:40		Kareem	OK
2	S1	S1	CAL2	10/14/24 13:45		Kareem	OK
3	S2	S2	CAL3	10/14/24 13:49		Kareem	OK
4	S3	S3	CAL4	10/14/24 13:53		Kareem	OK
5	S4	S4	CAL5	10/14/24 13:57		Kareem	OK
6	S5	S5	CAL6	10/14/24 14:02		Kareem	OK
7	ICV01	ICV01	ICV	10/14/24 14:06		Kareem	OK
8	LLICV01	LLICV01	LLICV	10/14/24 14:18		Kareem	OK
9	ICB01	ICB01	ICB	10/14/24 14:23		Kareem	OK
10	CRI01	CRI01	CRDL	10/14/24 14:27		Kareem	OK
11	ICSA01	ICSA01	ICSA	10/14/24 14:32		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	10/14/24 14:39		Kareem	OK
13	CCV01	CCV01	CCV	10/14/24 14:43		Kareem	OK
14	CCB01	CCB01	CCB	10/14/24 14:47		Kareem	OK
15	LR1	LR1	HIGH STD	10/14/24 14:53		Kareem	OK
16	LR2	LR2	HIGH STD	10/14/24 14:58		Kareem	OK
17	P4258-03	Chamberlain Ave	SAM	10/14/24 15:05		Kareem	OK
18	CCV02	CCV02	CCV	10/14/24 15:09		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132927

Review By	jaswal	Review On	10/21/2024 1:52:41 AM
Supervise By	mohan	Supervise On	10/21/2024 1:54:36 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441,MP82476,MP82477,MP82478,MP82479,MP82484		
ICV Standard	MP82485		
CCV Standard	MP82488		
ICSA Standard	MP82486,MP82487		
CRI Standard	MP82484		
LCS Standard			
Chk Standard	MP82491,MP82492		

19	CCB02	CCB02	CCB	10/14/24 15:13		Kareem	OK
20	P4354-01	291551	SAM	10/14/24 15:19		Kareem	OK
21	P4354-01DUP	291551DUP	DUP	10/14/24 15:24		Kareem	OK
22	P4354-01L	291551L	SD	10/14/24 15:28		Kareem	OK
23	P4354-01MS	291551MS	MS	10/14/24 15:32		Kareem	OK
24	P4354-01MSD	291551MSD	MSD	10/14/24 15:37		Kareem	OK
25	P4354-01A	291551A	PS	10/14/24 15:41		Kareem	OK
26	P4375-02	CONCRETE-PADS-1	SAM	10/14/24 15:54		Kareem	OK
27	P4375-04	CONCRETE-PADS-2	SAM	10/14/24 15:58		Kareem	OK
28	CCV03	CCV03	CCV	10/14/24 16:03		Kareem	OK
29	CCB03	CCB03	CCB	10/14/24 16:07		Kareem	OK
30	P4375-06	CONCRETE-PADS-3	SAM	10/14/24 16:21		Kareem	OK
31	P4375-08	CONCRETE-PADS-4	SAM	10/14/24 16:26		Kareem	OK
32	P4375-10	CONCRETE-PADS-5	SAM	10/14/24 16:30		Kareem	OK
33	P4375-12	CONCRETE-PADS-6	SAM	10/14/24 16:35		Kareem	OK
34	P4375-14	CONCRETE-PADS-7	SAM	10/14/24 16:39		Kareem	OK
35	P4375-16	CONCRETE-PADS-8	SAM	10/14/24 16:44		Kareem	OK
36	P4375-18	CONCRETE-PADS-9	SAM	10/14/24 16:48		Kareem	OK
37	P4375-20	CONCRETE-PADS-10	SAM	10/14/24 16:53		Kareem	OK
38	P4375-22	CONCRETE-PADS-11	SAM	10/14/24 16:57		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132927

Review By	jaswal	Review On	10/21/2024 1:52:41 AM
Supervise By	mohan	Supervise On	10/21/2024 1:54:36 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441,MP82476,MP82477,MP82478,MP82479,MP82484		
ICV Standard	MP82485		
CCV Standard	MP82488		
ICSA Standard	MP82486,MP82487		
CRI Standard	MP82484		
LCS Standard			
Chk Standard	MP82491,MP82492		

39	P4375-24	FENCE-POST-1	SAM	10/14/24 17:02		Kareem	OK
40	CCV04	CCV04	CCV	10/14/24 17:06		Kareem	OK
41	CCB04	CCB04	CCB	10/14/24 17:11		Kareem	OK
42	P4375-26	FENCE-POST-2	SAM	10/14/24 17:15		Kareem	OK
43	P4375-28	FENCE-POST-3	SAM	10/14/24 17:21		Kareem	OK
44	P4375-28DUP	FENCE-POST-3DUP	DUP	10/14/24 17:26		Kareem	OK
45	P4375-28L	FENCE-POST-3L	SD	10/14/24 17:30		Kareem	OK
46	P4375-28MS	FENCE-POST-3MS	MS	10/14/24 17:34		Kareem	OK
47	P4375-28MSD	FENCE-POST-3MSD	MSD	10/14/24 17:39		Kareem	OK
48	P4375-28A	FENCE-POST-3A	PS	10/14/24 17:43		Kareem	OK
49	PB164021TB	PB164021TB	MB	10/14/24 17:47		Kareem	OK
50	PB164077BL	PB164077BL	MB	10/14/24 17:52		Kareem	OK
51	PB164077BS	PB164077BS	LCS	10/14/24 17:56		Kareem	OK
52	CCV05	CCV05	CCV	10/14/24 18:00		Kareem	OK
53	CCB05	CCB05	CCB	10/14/24 18:04		Kareem	OK
54	PB164010TB	PB164010TB	MB	10/14/24 18:08		Kareem	OK
55	P4393-01	BUST-DEBRIS	SAM	10/14/24 18:13		Kareem	OK
56	P4393-01DUP	BUST-DEBRISDUP	DUP	10/14/24 18:17		Kareem	OK
57	P4393-01L	BUST-DEBRISL	SD	10/14/24 18:22		Kareem	OK
58	P4393-01MS	BUST-DEBRISMS	MS	10/14/24 18:26		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132927

Review By	jaswal	Review On	10/21/2024 1:52:41 AM
Supervise By	mohan	Supervise On	10/21/2024 1:54:36 AM

STD. NAME	STD REF.#
ICAL Standard	MP82441,MP82476,MP82477,MP82478,MP82479,MP82484
ICV Standard	MP82485
CCV Standard	MP82488
ICSA Standard	MP82486,MP82487
CRI Standard	MP82484
LCS Standard	
Chk Standard	MP82491,MP82492

59	P4393-01MSD	BUST-DEBRISMSD	MSD	10/14/24 18:30		Kareem	OK
60	P4393-01A	BUST-DEBRISA	PS	10/14/24 18:34		Kareem	OK
61	P4394-01	WASH-DEBRIS	SAM	10/14/24 18:38		Kareem	OK
62	P4395-01	F05308-SOLID	SAM	10/14/24 18:42		Kareem	OK
63	PB164106BL	PB164106BL	MB	10/14/24 18:46		Kareem	OK
64	CCV06	CCV06	CCV	10/14/24 18:51		Kareem	OK
65	CCB06	CCB06	CCB	10/14/24 18:55		Kareem	OK
66	PB164106BS	PB164106BS	LCS	10/14/24 18:59		Kareem	OK
67	PB163999BL	PB163999BL	MB	10/14/24 19:03		Kareem	OK
68	PB163999BS	PB163999BS	LCS	10/14/24 19:07		Kareem	OK
69	CCV07	CCV07	CCV	10/14/24 19:11		Kareem	OK
70	CCB07	CCB07	CCB	10/14/24 19:15		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132951

Review By	jaswal	Review On	10/21/2024 1:52:32 AM
Supervise By	mohan	Supervise On	10/21/2024 1:54:55 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441,MP82476,MP82477,MP82478,MP82479,MP82484		
ICV Standard	MP82485		
CCV Standard	MP82488		
ICSA Standard	MP82486,MP82487		
CRI Standard	MP82484		
LCS Standard			
Chk Standard	MP82491,MP82492		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/15/24 10:42		Kareem	OK
2	S1	S1	CAL2	10/15/24 10:47		Kareem	OK
3	S2	S2	CAL3	10/15/24 10:51		Kareem	OK
4	S3	S3	CAL4	10/15/24 10:55		Kareem	OK
5	S4	S4	CAL5	10/15/24 10:59		Kareem	OK
6	S5	S5	CAL6	10/15/24 11:03		Kareem	OK
7	ICV01	ICV01	ICV	10/15/24 11:24		Kareem	OK
8	LLICV01	LLICV01	LLICV	10/15/24 11:37		Kareem	OK
9	ICB01	ICB01	ICB	10/15/24 11:41		Kareem	OK
10	CRI01	CRI01	CRDL	10/15/24 11:46		Kareem	OK
11	ICSA01	ICSA01	ICSA	10/15/24 11:50		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	10/15/24 12:03		Kareem	OK
13	ICSADL	ICSADL	ICSA	10/15/24 12:10		Kareem	OK
14	ICSABDL	ICSABDL	ICSAB	10/15/24 12:20		Kareem	OK
15	CCV01	CCV01	CCV	10/15/24 12:26		Kareem	OK
16	CCB01	CCB01	CCB	10/15/24 12:30		Kareem	OK
17	LR1	LR1	HIGH STD	10/15/24 12:34		Kareem	OK
18	LR2	LR2	HIGH STD	10/15/24 12:39		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132951

Review By	jaswal	Review On	10/21/2024 1:52:32 AM
Supervise By	mohan	Supervise On	10/21/2024 1:54:55 AM

STD. NAME	STD REF.#
ICAL Standard	MP82441,MP82476,MP82477,MP82478,MP82479,MP82484
ICV Standard	MP82485
CCV Standard	MP82488
ICSA Standard	MP82486,MP82487
CRI Standard	MP82484
LCS Standard	
Chk Standard	MP82491,MP82492

19	P4271-07DL	14B-(1-4)-COMPDL	SAM	10/15/24 12:52	NOT USE	Kareem	Not Ok
20	P4271-07DL2	14B-(1-4)-COMPDL2	SAM	10/15/24 13:02	10x for Mn	Kareem	Confirms
21	PB164065BS	PB164065BS	LCS	10/15/24 13:06		Kareem	OK
22	P4385-18	SP-9	SAM	10/15/24 13:10		Kareem	OK
23	P4385-20	SP-10	SAM	10/15/24 13:14	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
24	P4385-20DUP	SP-10DUP	DUP	10/15/24 13:18	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
25	CCV02	CCV02	CCV	10/15/24 13:23		Kareem	OK
26	CCB02	CCB02	CCB	10/15/24 13:27		Kareem	OK
27	P4385-20L	SP-10L	SD	10/15/24 13:31	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
28	P4385-20MS	SP-10MS	MS	10/15/24 13:35	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
29	P4385-20MSD	SP-10MSD	MSD	10/15/24 13:39	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
30	P4385-20A	SP-10A	PS	10/15/24 13:43	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
31	CCV03	CCV03	CCV	10/15/24 14:07		Kareem	OK
32	CCB03	CCB03	CCB	10/15/24 14:11		Kareem	OK
33	PB164065BL	PB164065BL	MB	10/15/24 14:26		Kareem	OK
34	P4382-01	DECON-DRUM	SAM	10/15/24 14:30		Kareem	OK
35	CCV04	CCV04	CCV	10/15/24 14:35		Kareem	OK
36	CCB04	CCB04	CCB	10/15/24 14:39		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132951

Review By	jaswal	Review On	10/21/2024 1:52:32 AM
Supervise By	mohan	Supervise On	10/21/2024 1:54:55 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441,MP82476,MP82477,MP82478,MP82479,MP82484		
ICV Standard	MP82485		
CCV Standard	MP82488		
ICSA Standard	MP82486,MP82487		
CRI Standard	MP82484		
LCS Standard			
Chk Standard	MP82491,MP82492		

37	P4385-02	SP-1	SAM	10/15/24 14:43		Kareem	OK
38	P4385-04	SP-2	SAM	10/15/24 14:48		Kareem	OK
39	P4385-06	SP-3	SAM	10/15/24 14:52		Kareem	OK
40	P4385-08	SP-4	SAM	10/15/24 14:56		Kareem	OK
41	P4385-10	SP-5	SAM	10/15/24 15:01		Kareem	OK
42	P4385-12	SP-6	SAM	10/15/24 15:05		Kareem	OK
43	P4385-14	SP-7	SAM	10/15/24 15:09		Kareem	OK
44	P4385-16	SP-8	SAM	10/15/24 15:13		Kareem	OK
45	P4370-02	COMP	SAM	10/15/24 15:18		Kareem	OK
46	P4396-01	WASTE-WATER-FRA	SAM	10/15/24 15:22		Kareem	OK
47	CCV05	CCV05	CCV	10/15/24 15:26		Kareem	OK
48	CCB05	CCB05	CCB	10/15/24 15:30		Kareem	OK
49	P4396-01DUP	WASTE-WATER-FRA	DUP	10/15/24 15:35		Kareem	OK
50	P4396-01L	WASTE-WATER-FRA	SD	10/15/24 15:39		Kareem	OK
51	P4396-01MS	WASTE-WATER-FRA	MS	10/15/24 15:43		Kareem	OK
52	P4396-01MSD	WASTE-WATER-FRA	MSD	10/15/24 15:47		Kareem	OK
53	P4396-01A	WASTE-WATER-FRA	PS	10/15/24 15:51		Kareem	OK
54	PB164099BL	PB164099BL	MB	10/15/24 15:55		Kareem	OK
55	PB164099BS	PB164099BS	LCS	10/15/24 15:59		Kareem	OK
56	P4364-01	SB-01-1.0-1.5	SAM	10/15/24 16:03		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132951

Review By	jaswal	Review On	10/21/2024 1:52:32 AM
Supervise By	mohan	Supervise On	10/21/2024 1:54:55 AM

STD. NAME	STD REF.#
ICAL Standard	MP82441,MP82476,MP82477,MP82478,MP82479,MP82484
ICV Standard	MP82485
CCV Standard	MP82488
ICSA Standard	MP82486,MP82487
CRI Standard	MP82484
LCS Standard	
Chk Standard	MP82491,MP82492

57	P4364-02	SB-01-2.5-3.0	SAM	10/15/24 16:08		Kareem	OK
58	P4364-03	SB-01-4.25-4.83	SAM	10/15/24 16:13		Kareem	OK
59	CCV06	CCV06	CCV	10/15/24 16:17		Kareem	OK
60	CCB06	CCB06	CCB	10/15/24 16:21		Kareem	OK
61	P4400-01	NB-08-101424	SAM	10/15/24 16:26	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
62	P4400-01DUP	NB-08-101424DUP	DUP	10/15/24 16:30	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
63	P4400-01L	NB-08-101424L	SD	10/15/24 16:34	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
64	P4400-01MS	NB-08-101424MS	MS	10/15/24 16:39	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
65	P4400-01MSD	NB-08-101424MSD	MSD	10/15/24 16:43	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
66	P4400-01A	NB-08-101424A	PS	10/15/24 16:47	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
67	P4364-04	SB-02-0.167-0.667	SAM	10/15/24 16:52		Kareem	OK
68	P4364-05	SB-02-2.0-2.5	SAM	10/15/24 16:56		Kareem	OK
69	P4364-06	SB-02-4.5-5.0	SAM	10/15/24 17:01		Kareem	OK
70	P4364-07	SB-03-0.167-0.667	SAM	10/15/24 17:05		Kareem	OK
71	CCV07	CCV07	CCV	10/15/24 17:14		Kareem	OK
72	CCB07	CCB07	CCB	10/15/24 17:18		Kareem	OK
73	P4364-08	SB-03-2.0-2.5	SAM	10/15/24 17:23		Kareem	OK
74	P4364-09	SB-03-4.167-4.667	SAM	10/15/24 17:27		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132951

Review By	jaswal	Review On	10/21/2024 1:52:32 AM
Supervise By	mohan	Supervise On	10/21/2024 1:54:55 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441,MP82476,MP82477,MP82478,MP82479,MP82484		
ICV Standard	MP82485		
CCV Standard	MP82488		
ICSA Standard	MP82486,MP82487		
CRI Standard	MP82484		
LCS Standard			
Chk Standard	MP82491,MP82492		

75	P4364-10	SB-04-0.167-0.667	SAM	10/15/24 17:31		Kareem	OK
76	P4364-11	SB-04-2.0-2.5	SAM	10/15/24 17:36	Not used	Kareem	Not Ok
77	P4364-12	SB-04-4.5-5.0	SAM	10/15/24 17:41		Kareem	OK
78	P4364-13	SB-05-0.167-0.667	SAM	10/15/24 17:45	Ca high	Kareem	Dilution
79	CCV08	CCV08	CCV	10/15/24 17:49		Kareem	OK
80	CCB08	CCB08	CCB	10/15/24 17:53		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133014

Review By	jaswal	Review On	10/21/2024 1:53:28 AM
Supervise By	mohan	Supervise On	10/21/2024 1:54:16 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441,MP82476,MP82477,MP82478,MP82479,MP82484		
ICV Standard	MP82485		
CCV Standard	MP82488		
ICSA Standard	MP82486,MP82487		
CRI Standard	MP82484		
LCS Standard			
Chk Standard	MP82491,MP82492		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/18/24 12:03		Kareem	OK
2	S1	S1	CAL2	10/18/24 12:07		Kareem	OK
3	S2	S2	CAL3	10/18/24 12:12		Kareem	OK
4	S3	S3	CAL4	10/18/24 12:16		Kareem	OK
5	S4	S4	CAL5	10/18/24 12:20		Kareem	OK
6	S5	S5	CAL6	10/18/24 12:24		Kareem	OK
7	ICV01	ICV01	ICV	10/18/24 13:08	ICV fail for Al,Cr,Cu,Ag,Tl,Zn (200.7)	Kareem	OK
8	LLICV01	LLICV01	LLICV	10/18/24 13:13		Kareem	OK
9	ICB01	ICB01	ICB	10/18/24 13:23		Kareem	OK
10	CRI01	CRI01	CRDL	10/18/24 13:27		Kareem	OK
11	ICSA01	ICSA01	ICSA	10/18/24 13:32		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	10/18/24 13:56		Kareem	OK
13	CCV01	CCV01	CCV	10/18/24 14:00		Kareem	OK
14	CCB01	CCB01	CCB	10/18/24 14:05		Kareem	OK
15	P3846-05DL	PT-MIN2-WPDL	SAM	10/18/24 14:19	2x for Ca,Na	Kareem	Confirms
16	P3847-05DL	PT-MIN2-WPDL	SAM	10/18/24 14:49	2x for Ca,Na	Kareem	Confirms
17	CCV02	CCV02	CCV	10/18/24 14:53		Kareem	OK
18	CCB02	CCB02	CCB	10/18/24 14:57		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133014

Review By	jaswal	Review On	10/21/2024 1:53:28 AM
Supervise By	mohan	Supervise On	10/21/2024 1:54:16 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441,MP82476,MP82477,MP82478,MP82479,MP82484		
ICV Standard	MP82485		
CCV Standard	MP82488		
ICSA Standard	MP82486,MP82487		
CRI Standard	MP82484		
LCS Standard			
Chk Standard	MP82491,MP82492		

19	LR1	LR1	HIGH STD	10/18/24 15:02		Kareem	OK
20	LR2	LR2	HIGH STD	10/18/24 15:07		Kareem	OK
21	P4385-20	SP-10	SAM	10/18/24 15:11		Kareem	OK
22	P4385-20DUP	SP-10DUP	DUP	10/18/24 15:16		Kareem	OK
23	P4385-20L	SP-10L	SD	10/18/24 15:20		Kareem	OK
24	P4385-20MS	SP-10MS	MS	10/18/24 15:24		Kareem	OK
25	P4385-20MSD	SP-10MSD	MSD	10/18/24 15:28		Kareem	OK
26	P4385-20A	SP-10A	PS	10/18/24 15:32		Kareem	OK
27	CCV03	CCV03	CCV	10/18/24 16:05		Kareem	OK
28	CCB03	CCB03	CCB	10/18/24 16:39		Kareem	OK
29	P4405-01DL	MH-121DL	SAM	10/18/24 16:43	5x for Na	Kareem	Confirms
30	P4421-01	EO-02-101624	SAM	10/18/24 17:15	NOT USE	Kareem	Not Ok
31	P4422-01	EO-01-101624	SAM	10/18/24 17:19	NOT USE	Kareem	Not Ok
32	P4431-01	72-11934	SAM	10/18/24 17:23	NOT USE	Kareem	Not Ok
33	CCV04	CCV04	CCV	10/18/24 17:28	CCV fail for Fe,Mg,Mn	Kareem	OK
34	CCB04	CCB04	CCB	10/18/24 17:32		Kareem	OK