



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Roman E&G Corp				SDG No.:	P4258			
Contract:	ROMA02	Lab Code:	CHEM		Case No.:	P4258	SAS No.:	P4258	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB36	Mercury	0.20	+/-0.20	U	0.20	CV	10/14/2024	11:38	LB132914

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Roman E&G Corp</u>		SDG No.: <u>P4258</u>							
Contract: <u>ROMA02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4258</u>	SAS No.: <u>P4258</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB13	Mercury	0.20	+/-0.20	U	0.20	CV	10/14/2024	11:43	LB132914
CCB14	Mercury	0.20	+/-0.20	U	0.20	CV	10/14/2024	12:13	LB132914
CCB15	Mercury	0.20	+/-0.20	U	0.20	CV	10/14/2024	12:41	LB132914
CCB16	Mercury	0.20	+/-0.20	U	0.20	CV	10/14/2024	13:12	LB132914
CCB17	Mercury	0.20	+/-0.20	U	0.20	CV	10/14/2024	13:39	LB132914
CCB18	Mercury	0.20	+/-0.20	U	0.20	CV	10/14/2024	14:06	LB132914
CCB19	Mercury	0.20	+/-0.20	U	0.20	CV	10/14/2024	14:34	LB132914
CCB20	Mercury	0.20	+/-0.20	U	0.20	CV	10/14/2024	15:01	LB132914
CCB21	Mercury	0.20	+/-0.20	U	0.20	CV	10/14/2024	15:17	LB132914

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Roman E&G Corp</u>		SDG No.: <u>P4258</u>							
Contract: <u>ROMA02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4258</u>	SAS No.: <u>P4258</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	10/14/2024	14:23	LB132927
	Antimony	50.0	+/-50.0	U	50.0	P	10/14/2024	14:23	LB132927
	Arsenic	20.0	+/-20.0	U	20.0	P	10/14/2024	14:23	LB132927
	Barium	100	+/-100	U	100	P	10/14/2024	14:23	LB132927
	Beryllium	6.00	+/-6.00	U	6.00	P	10/14/2024	14:23	LB132927
	Cadmium	6.00	+/-6.00	U	6.00	P	10/14/2024	14:23	LB132927
	Calcium	2000	+/-2000	U	2000	P	10/14/2024	14:23	LB132927
	Chromium	10.0	+/-10.0	U	10.0	P	10/14/2024	14:23	LB132927
	Cobalt	30.0	+/-30.0	U	30.0	P	10/14/2024	14:23	LB132927
	Copper	20.0	+/-20.0	U	20.0	P	10/14/2024	14:23	LB132927
	Iron	100	+/-100	U	100	P	10/14/2024	14:23	LB132927
	Lead	12.0	+/-12.0	U	12.0	P	10/14/2024	14:23	LB132927
	Magnesium	2000	+/-2000	U	2000	P	10/14/2024	14:23	LB132927
	Manganese	20.0	+/-20.0	U	20.0	P	10/14/2024	14:23	LB132927
	Nickel	40.0	+/-40.0	U	40.0	P	10/14/2024	14:23	LB132927
	Potassium	2000	+/-2000	U	2000	P	10/14/2024	14:23	LB132927
	Selenium	20.0	+/-20.0	U	20.0	P	10/14/2024	14:23	LB132927
	Silver	10.0	+/-10.0	U	10.0	P	10/14/2024	14:23	LB132927
	Sodium	2000	+/-2000	U	2000	P	10/14/2024	14:23	LB132927
	Thallium	40.0	+/-40.0	U	40.0	P	10/14/2024	14:23	LB132927
	Vanadium	40.0	+/-40.0	U	40.0	P	10/14/2024	14:23	LB132927
	Zinc	40.0	+/-40.0	U	40.0	P	10/14/2024	14:23	LB132927

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Roman E&G Corp **SDG No.:** P4258
Contract: ROMA02 **Lab Code:** CHEM **Case No.:** P4258 **SAS No.:** P4258

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	10/14/2024	14:47	LB132927
	Antimony	50.0	+/-50.0	U	50.0	P	10/14/2024	14:47	LB132927
	Arsenic	20.0	+/-20.0	U	20.0	P	10/14/2024	14:47	LB132927
	Barium	100	+/-100	U	100	P	10/14/2024	14:47	LB132927
	Beryllium	6.00	+/-6.00	U	6.00	P	10/14/2024	14:47	LB132927
	Cadmium	6.00	+/-6.00	U	6.00	P	10/14/2024	14:47	LB132927
	Calcium	2000	+/-2000	U	2000	P	10/14/2024	14:47	LB132927
	Chromium	10.0	+/-10.0	U	10.0	P	10/14/2024	14:47	LB132927
	Cobalt	30.0	+/-30.0	U	30.0	P	10/14/2024	14:47	LB132927
	Copper	20.0	+/-20.0	U	20.0	P	10/14/2024	14:47	LB132927
	Iron	100	+/-100	U	100	P	10/14/2024	14:47	LB132927
	Lead	12.0	+/-12.0	U	12.0	P	10/14/2024	14:47	LB132927
	Magnesium	2000	+/-2000	U	2000	P	10/14/2024	14:47	LB132927
	Manganese	20.0	+/-20.0	U	20.0	P	10/14/2024	14:47	LB132927
	Nickel	40.0	+/-40.0	U	40.0	P	10/14/2024	14:47	LB132927
	Potassium	2000	+/-2000	U	2000	P	10/14/2024	14:47	LB132927
	Selenium	20.0	+/-20.0	U	20.0	P	10/14/2024	14:47	LB132927
	Silver	10.0	+/-10.0	U	10.0	P	10/14/2024	14:47	LB132927
	Sodium	2000	+/-2000	U	2000	P	10/14/2024	14:47	LB132927
	Thallium	40.0	+/-40.0	U	40.0	P	10/14/2024	14:47	LB132927
	Vanadium	40.0	+/-40.0	U	40.0	P	10/14/2024	14:47	LB132927
	Zinc	40.0	+/-40.0	U	40.0	P	10/14/2024	14:47	LB132927
CCB02	Aluminum	100	+/-100	U	100	P	10/14/2024	15:13	LB132927
	Antimony	50.0	+/-50.0	U	50.0	P	10/14/2024	15:13	LB132927
	Arsenic	20.0	+/-20.0	U	20.0	P	10/14/2024	15:13	LB132927
	Barium	100	+/-100	U	100	P	10/14/2024	15:13	LB132927
	Beryllium	6.00	+/-6.00	U	6.00	P	10/14/2024	15:13	LB132927
	Cadmium	6.00	+/-6.00	U	6.00	P	10/14/2024	15:13	LB132927
	Calcium	2000	+/-2000	U	2000	P	10/14/2024	15:13	LB132927
	Chromium	10.0	+/-10.0	U	10.0	P	10/14/2024	15:13	LB132927
	Cobalt	30.0	+/-30.0	U	30.0	P	10/14/2024	15:13	LB132927
	Copper	20.0	+/-20.0	U	20.0	P	10/14/2024	15:13	LB132927
	Iron	100	+/-100	U	100	P	10/14/2024	15:13	LB132927
	Lead	12.0	+/-12.0	U	12.0	P	10/14/2024	15:13	LB132927
	Magnesium	2000	+/-2000	U	2000	P	10/14/2024	15:13	LB132927
	Manganese	20.0	+/-20.0	U	20.0	P	10/14/2024	15:13	LB132927
	Nickel	40.0	+/-40.0	U	40.0	P	10/14/2024	15:13	LB132927
	Potassium	2000	+/-2000	U	2000	P	10/14/2024	15:13	LB132927
	Selenium	20.0	+/-20.0	U	20.0	P	10/14/2024	15:13	LB132927

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Roman E&G Corp		SDG No.: P4258							
Contract: ROMA02	Lab Code: CHEM	Case No.: P4258	SAS No.: P4258						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	10/14/2024	15:13	LB132927
	Sodium	2000	+/-2000	U	2000	P	10/14/2024	15:13	LB132927
	Thallium	40.0	+/-40.0	U	40.0	P	10/14/2024	15:13	LB132927
	Vanadium	40.0	+/-40.0	U	40.0	P	10/14/2024	15:13	LB132927
	Zinc	40.0	+/-40.0	U	40.0	P	10/14/2024	15:13	LB132927
CCB03	Aluminum	100	+/-100	U	100	P	10/14/2024	16:07	LB132927
	Antimony	50.0	+/-50.0	U	50.0	P	10/14/2024	16:07	LB132927
	Arsenic	20.0	+/-20.0	U	20.0	P	10/14/2024	16:07	LB132927
	Barium	100	+/-100	U	100	P	10/14/2024	16:07	LB132927
	Beryllium	6.00	+/-6.00	U	6.00	P	10/14/2024	16:07	LB132927
	Cadmium	6.00	+/-6.00	U	6.00	P	10/14/2024	16:07	LB132927
	Calcium	2000	+/-2000	U	2000	P	10/14/2024	16:07	LB132927
	Chromium	10.0	+/-10.0	U	10.0	P	10/14/2024	16:07	LB132927
	Cobalt	30.0	+/-30.0	U	30.0	P	10/14/2024	16:07	LB132927
	Copper	20.0	+/-20.0	U	20.0	P	10/14/2024	16:07	LB132927
	Iron	100	+/-100	U	100	P	10/14/2024	16:07	LB132927
	Lead	12.0	+/-12.0	U	12.0	P	10/14/2024	16:07	LB132927
	Magnesium	2000	+/-2000	U	2000	P	10/14/2024	16:07	LB132927
	Manganese	20.0	+/-20.0	U	20.0	P	10/14/2024	16:07	LB132927
	Nickel	40.0	+/-40.0	U	40.0	P	10/14/2024	16:07	LB132927
	Potassium	2000	+/-2000	U	2000	P	10/14/2024	16:07	LB132927
	Selenium	20.0	+/-20.0	U	20.0	P	10/14/2024	16:07	LB132927
	Silver	1.36	+/-10.0	J	10.0	P	10/14/2024	16:07	LB132927
	Sodium	2000	+/-2000	U	2000	P	10/14/2024	16:07	LB132927
	Thallium	40.0	+/-40.0	U	40.0	P	10/14/2024	16:07	LB132927
	Vanadium	40.0	+/-40.0	U	40.0	P	10/14/2024	16:07	LB132927
	Zinc	40.0	+/-40.0	U	40.0	P	10/14/2024	16:07	LB132927
CCB04	Aluminum	100	+/-100	U	100	P	10/14/2024	17:11	LB132927
	Antimony	50.0	+/-50.0	U	50.0	P	10/14/2024	17:11	LB132927
	Arsenic	20.0	+/-20.0	U	20.0	P	10/14/2024	17:11	LB132927
	Barium	100	+/-100	U	100	P	10/14/2024	17:11	LB132927
	Beryllium	6.00	+/-6.00	U	6.00	P	10/14/2024	17:11	LB132927
	Cadmium	6.00	+/-6.00	U	6.00	P	10/14/2024	17:11	LB132927
	Calcium	2000	+/-2000	U	2000	P	10/14/2024	17:11	LB132927
	Chromium	10.0	+/-10.0	U	10.0	P	10/14/2024	17:11	LB132927
	Cobalt	30.0	+/-30.0	U	30.0	P	10/14/2024	17:11	LB132927
	Copper	20.0	+/-20.0	U	20.0	P	10/14/2024	17:11	LB132927
	Iron	100	+/-100	U	100	P	10/14/2024	17:11	LB132927
	Lead	12.0	+/-12.0	U	12.0	P	10/14/2024	17:11	LB132927

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Roman E&G Corp		SDG No.: P4258							
Contract: ROMA02	Lab Code: CHEM	Case No.: P4258	SAS No.: P4258						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	10/14/2024	17:11	LB132927
	Manganese	20.0	+/-20.0	U	20.0	P	10/14/2024	17:11	LB132927
	Nickel	40.0	+/-40.0	U	40.0	P	10/14/2024	17:11	LB132927
	Potassium	2000	+/-2000	U	2000	P	10/14/2024	17:11	LB132927
	Selenium	20.0	+/-20.0	U	20.0	P	10/14/2024	17:11	LB132927
	Silver	1.36	+/-10.0	J	10.0	P	10/14/2024	17:11	LB132927
	Sodium	2000	+/-2000	U	2000	P	10/14/2024	17:11	LB132927
	Thallium	40.0	+/-40.0	U	40.0	P	10/14/2024	17:11	LB132927
	Vanadium	40.0	+/-40.0	U	40.0	P	10/14/2024	17:11	LB132927
	Zinc	40.0	+/-40.0	U	40.0	P	10/14/2024	17:11	LB132927
CCB05	Aluminum	100	+/-100	U	100	P	10/14/2024	18:04	LB132927
	Antimony	50.0	+/-50.0	U	50.0	P	10/14/2024	18:04	LB132927
	Arsenic	20.0	+/-20.0	U	20.0	P	10/14/2024	18:04	LB132927
	Barium	100	+/-100	U	100	P	10/14/2024	18:04	LB132927
	Beryllium	6.00	+/-6.00	U	6.00	P	10/14/2024	18:04	LB132927
	Cadmium	6.00	+/-6.00	U	6.00	P	10/14/2024	18:04	LB132927
	Calcium	2000	+/-2000	U	2000	P	10/14/2024	18:04	LB132927
	Chromium	10.0	+/-10.0	U	10.0	P	10/14/2024	18:04	LB132927
	Cobalt	30.0	+/-30.0	U	30.0	P	10/14/2024	18:04	LB132927
	Copper	20.0	+/-20.0	U	20.0	P	10/14/2024	18:04	LB132927
	Iron	100	+/-100	U	100	P	10/14/2024	18:04	LB132927
	Lead	12.0	+/-12.0	U	12.0	P	10/14/2024	18:04	LB132927
	Magnesium	2000	+/-2000	U	2000	P	10/14/2024	18:04	LB132927
	Manganese	20.0	+/-20.0	U	20.0	P	10/14/2024	18:04	LB132927
	Nickel	40.0	+/-40.0	U	40.0	P	10/14/2024	18:04	LB132927
	Potassium	2000	+/-2000	U	2000	P	10/14/2024	18:04	LB132927
	Selenium	20.0	+/-20.0	U	20.0	P	10/14/2024	18:04	LB132927
	Silver	1.27	+/-10.0	J	10.0	P	10/14/2024	18:04	LB132927
	Sodium	2000	+/-2000	U	2000	P	10/14/2024	18:04	LB132927
	Thallium	40.0	+/-40.0	U	40.0	P	10/14/2024	18:04	LB132927
	Vanadium	40.0	+/-40.0	U	40.0	P	10/14/2024	18:04	LB132927
	Zinc	40.0	+/-40.0	U	40.0	P	10/14/2024	18:04	LB132927
CCB06	Aluminum	100	+/-100	U	100	P	10/14/2024	18:55	LB132927
	Antimony	50.0	+/-50.0	U	50.0	P	10/14/2024	18:55	LB132927
	Arsenic	20.0	+/-20.0	U	20.0	P	10/14/2024	18:55	LB132927
	Barium	100	+/-100	U	100	P	10/14/2024	18:55	LB132927
	Beryllium	6.00	+/-6.00	U	6.00	P	10/14/2024	18:55	LB132927
	Cadmium	6.00	+/-6.00	U	6.00	P	10/14/2024	18:55	LB132927
	Calcium	2000	+/-2000	U	2000	P	10/14/2024	18:55	LB132927

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Roman E&G Corp		SDG No.: P4258							
Contract: ROMA02	Lab Code: CHEM	Case No.: P4258	SAS No.: P4258						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	10/14/2024	18:55	LB132927
	Cobalt	30.0	+/-30.0	U	30.0	P	10/14/2024	18:55	LB132927
	Copper	20.0	+/-20.0	U	20.0	P	10/14/2024	18:55	LB132927
	Iron	100	+/-100	U	100	P	10/14/2024	18:55	LB132927
	Lead	12.0	+/-12.0	U	12.0	P	10/14/2024	18:55	LB132927
	Magnesium	2000	+/-2000	U	2000	P	10/14/2024	18:55	LB132927
	Manganese	20.0	+/-20.0	U	20.0	P	10/14/2024	18:55	LB132927
	Nickel	40.0	+/-40.0	U	40.0	P	10/14/2024	18:55	LB132927
	Potassium	2000	+/-2000	U	2000	P	10/14/2024	18:55	LB132927
	Selenium	20.0	+/-20.0	U	20.0	P	10/14/2024	18:55	LB132927
	Silver	10.0	+/-10.0	U	10.0	P	10/14/2024	18:55	LB132927
	Sodium	2000	+/-2000	U	2000	P	10/14/2024	18:55	LB132927
	Thallium	40.0	+/-40.0	U	40.0	P	10/14/2024	18:55	LB132927
	Vanadium	40.0	+/-40.0	U	40.0	P	10/14/2024	18:55	LB132927
	Zinc	40.0	+/-40.0	U	40.0	P	10/14/2024	18:55	LB132927
CCB07	Aluminum	100	+/-100	U	100	P	10/14/2024	19:15	LB132927
	Antimony	50.0	+/-50.0	U	50.0	P	10/14/2024	19:15	LB132927
	Arsenic	20.0	+/-20.0	U	20.0	P	10/14/2024	19:15	LB132927
	Barium	100	+/-100	U	100	P	10/14/2024	19:15	LB132927
	Beryllium	6.00	+/-6.00	U	6.00	P	10/14/2024	19:15	LB132927
	Cadmium	6.00	+/-6.00	U	6.00	P	10/14/2024	19:15	LB132927
	Calcium	2000	+/-2000	U	2000	P	10/14/2024	19:15	LB132927
	Chromium	10.0	+/-10.0	U	10.0	P	10/14/2024	19:15	LB132927
	Cobalt	30.0	+/-30.0	U	30.0	P	10/14/2024	19:15	LB132927
	Copper	20.0	+/-20.0	U	20.0	P	10/14/2024	19:15	LB132927
	Iron	100	+/-100	U	100	P	10/14/2024	19:15	LB132927
	Lead	12.0	+/-12.0	U	12.0	P	10/14/2024	19:15	LB132927
	Magnesium	2000	+/-2000	U	2000	P	10/14/2024	19:15	LB132927
	Manganese	20.0	+/-20.0	U	20.0	P	10/14/2024	19:15	LB132927
	Nickel	40.0	+/-40.0	U	40.0	P	10/14/2024	19:15	LB132927
	Potassium	2000	+/-2000	U	2000	P	10/14/2024	19:15	LB132927
	Selenium	20.0	+/-20.0	U	20.0	P	10/14/2024	19:15	LB132927
	Silver	1.22	+/-10.0	J	10.0	P	10/14/2024	19:15	LB132927
	Sodium	2000	+/-2000	U	2000	P	10/14/2024	19:15	LB132927
	Thallium	40.0	+/-40.0	U	40.0	P	10/14/2024	19:15	LB132927
	Vanadium	40.0	+/-40.0	U	40.0	P	10/14/2024	19:15	LB132927
	Zinc	40.0	+/-40.0	U	40.0	P	10/14/2024	19:15	LB132927

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Roman E&G Corp</u>		SDG No.: <u>P4258</u>							
Contract: <u>ROMA02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4258</u>	SAS No.: <u>P4258</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	10/15/2024	11:41	LB132951
	Antimony	50.0	+/-50.0	U	50.0	P	10/15/2024	11:41	LB132951
	Arsenic	20.0	+/-20.0	U	20.0	P	10/15/2024	11:41	LB132951
	Barium	100	+/-100	U	100	P	10/15/2024	11:41	LB132951
	Beryllium	6.00	+/-6.00	U	6.00	P	10/15/2024	11:41	LB132951
	Cadmium	6.00	+/-6.00	U	6.00	P	10/15/2024	11:41	LB132951
	Calcium	2000	+/-2000	U	2000	P	10/15/2024	11:41	LB132951
	Chromium	10.0	+/-10.0	U	10.0	P	10/15/2024	11:41	LB132951
	Cobalt	30.0	+/-30.0	U	30.0	P	10/15/2024	11:41	LB132951
	Copper	20.0	+/-20.0	U	20.0	P	10/15/2024	11:41	LB132951
	Iron	100	+/-100	U	100	P	10/15/2024	11:41	LB132951
	Lead	12.0	+/-12.0	U	12.0	P	10/15/2024	11:41	LB132951
	Magnesium	2000	+/-2000	U	2000	P	10/15/2024	11:41	LB132951
	Manganese	20.0	+/-20.0	U	20.0	P	10/15/2024	11:41	LB132951
	Nickel	40.0	+/-40.0	U	40.0	P	10/15/2024	11:41	LB132951
	Potassium	2000	+/-2000	U	2000	P	10/15/2024	11:41	LB132951
	Selenium	20.0	+/-20.0	U	20.0	P	10/15/2024	11:41	LB132951
	Silver	10.0	+/-10.0	U	10.0	P	10/15/2024	11:41	LB132951
	Sodium	2000	+/-2000	U	2000	P	10/15/2024	11:41	LB132951
	Thallium	40.0	+/-40.0	U	40.0	P	10/15/2024	11:41	LB132951
	Vanadium	40.0	+/-40.0	U	40.0	P	10/15/2024	11:41	LB132951
	Zinc	40.0	+/-40.0	U	40.0	P	10/15/2024	11:41	LB132951

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Roman E&G Corp **SDG No.:** P4258
Contract: ROMA02 **Lab Code:** CHEM **Case No.:** P4258 **SAS No.:** P4258

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	10/15/2024	12:30	LB132951
	Antimony	50.0	+/-50.0	U	50.0	P	10/15/2024	12:30	LB132951
	Arsenic	20.0	+/-20.0	U	20.0	P	10/15/2024	12:30	LB132951
	Barium	100	+/-100	U	100	P	10/15/2024	12:30	LB132951
	Beryllium	6.00	+/-6.00	U	6.00	P	10/15/2024	12:30	LB132951
	Cadmium	6.00	+/-6.00	U	6.00	P	10/15/2024	12:30	LB132951
	Calcium	2000	+/-2000	U	2000	P	10/15/2024	12:30	LB132951
	Chromium	10.0	+/-10.0	U	10.0	P	10/15/2024	12:30	LB132951
	Cobalt	30.0	+/-30.0	U	30.0	P	10/15/2024	12:30	LB132951
	Copper	20.0	+/-20.0	U	20.0	P	10/15/2024	12:30	LB132951
	Iron	100	+/-100	U	100	P	10/15/2024	12:30	LB132951
	Lead	12.0	+/-12.0	U	12.0	P	10/15/2024	12:30	LB132951
	Magnesium	2000	+/-2000	U	2000	P	10/15/2024	12:30	LB132951
	Manganese	20.0	+/-20.0	U	20.0	P	10/15/2024	12:30	LB132951
	Nickel	40.0	+/-40.0	U	40.0	P	10/15/2024	12:30	LB132951
	Potassium	2000	+/-2000	U	2000	P	10/15/2024	12:30	LB132951
	Selenium	20.0	+/-20.0	U	20.0	P	10/15/2024	12:30	LB132951
	Silver	10.0	+/-10.0	U	10.0	P	10/15/2024	12:30	LB132951
	Sodium	2000	+/-2000	U	2000	P	10/15/2024	12:30	LB132951
	Thallium	40.0	+/-40.0	U	40.0	P	10/15/2024	12:30	LB132951
	Vanadium	40.0	+/-40.0	U	40.0	P	10/15/2024	12:30	LB132951
	Zinc	40.0	+/-40.0	U	40.0	P	10/15/2024	12:30	LB132951
CCB02	Aluminum	100	+/-100	U	100	P	10/15/2024	13:27	LB132951
	Antimony	50.0	+/-50.0	U	50.0	P	10/15/2024	13:27	LB132951
	Arsenic	20.0	+/-20.0	U	20.0	P	10/15/2024	13:27	LB132951
	Barium	100	+/-100	U	100	P	10/15/2024	13:27	LB132951
	Beryllium	6.00	+/-6.00	U	6.00	P	10/15/2024	13:27	LB132951
	Cadmium	6.00	+/-6.00	U	6.00	P	10/15/2024	13:27	LB132951
	Calcium	2000	+/-2000	U	2000	P	10/15/2024	13:27	LB132951
	Chromium	10.0	+/-10.0	U	10.0	P	10/15/2024	13:27	LB132951
	Cobalt	30.0	+/-30.0	U	30.0	P	10/15/2024	13:27	LB132951
	Copper	20.0	+/-20.0	U	20.0	P	10/15/2024	13:27	LB132951
	Iron	100	+/-100	U	100	P	10/15/2024	13:27	LB132951
	Lead	12.0	+/-12.0	U	12.0	P	10/15/2024	13:27	LB132951
	Magnesium	2000	+/-2000	U	2000	P	10/15/2024	13:27	LB132951
	Manganese	20.0	+/-20.0	U	20.0	P	10/15/2024	13:27	LB132951
	Nickel	40.0	+/-40.0	U	40.0	P	10/15/2024	13:27	LB132951
	Potassium	2000	+/-2000	U	2000	P	10/15/2024	13:27	LB132951
	Selenium	20.0	+/-20.0	U	20.0	P	10/15/2024	13:27	LB132951

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Roman E&G Corp		SDG No.: P4258							
Contract: ROMA02	Lab Code: CHEM	Case No.: P4258	SAS No.: P4258						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	10/15/2024	13:27	LB132951
	Sodium	2000	+/-2000	U	2000	P	10/15/2024	13:27	LB132951
	Thallium	40.0	+/-40.0	U	40.0	P	10/15/2024	13:27	LB132951
	Vanadium	40.0	+/-40.0	U	40.0	P	10/15/2024	13:27	LB132951
	Zinc	40.0	+/-40.0	U	40.0	P	10/15/2024	13:27	LB132951
CCB03	Aluminum	100	+/-100	U	100	P	10/15/2024	14:11	LB132951
	Antimony	50.0	+/-50.0	U	50.0	P	10/15/2024	14:11	LB132951
	Arsenic	20.0	+/-20.0	U	20.0	P	10/15/2024	14:11	LB132951
	Barium	100	+/-100	U	100	P	10/15/2024	14:11	LB132951
	Beryllium	6.00	+/-6.00	U	6.00	P	10/15/2024	14:11	LB132951
	Cadmium	6.00	+/-6.00	U	6.00	P	10/15/2024	14:11	LB132951
	Calcium	2000	+/-2000	U	2000	P	10/15/2024	14:11	LB132951
	Chromium	10.0	+/-10.0	U	10.0	P	10/15/2024	14:11	LB132951
	Cobalt	30.0	+/-30.0	U	30.0	P	10/15/2024	14:11	LB132951
	Copper	20.0	+/-20.0	U	20.0	P	10/15/2024	14:11	LB132951
	Iron	100	+/-100	U	100	P	10/15/2024	14:11	LB132951
	Lead	12.0	+/-12.0	U	12.0	P	10/15/2024	14:11	LB132951
	Magnesium	2000	+/-2000	U	2000	P	10/15/2024	14:11	LB132951
	Manganese	20.0	+/-20.0	U	20.0	P	10/15/2024	14:11	LB132951
	Nickel	40.0	+/-40.0	U	40.0	P	10/15/2024	14:11	LB132951
	Potassium	2000	+/-2000	U	2000	P	10/15/2024	14:11	LB132951
	Selenium	20.0	+/-20.0	U	20.0	P	10/15/2024	14:11	LB132951
	Silver	10.0	+/-10.0	U	10.0	P	10/15/2024	14:11	LB132951
	Sodium	2000	+/-2000	U	2000	P	10/15/2024	14:11	LB132951
	Thallium	40.0	+/-40.0	U	40.0	P	10/15/2024	14:11	LB132951
	Vanadium	40.0	+/-40.0	U	40.0	P	10/15/2024	14:11	LB132951
	Zinc	40.0	+/-40.0	U	40.0	P	10/15/2024	14:11	LB132951
CCB04	Aluminum	100	+/-100	U	100	P	10/15/2024	14:39	LB132951
	Antimony	50.0	+/-50.0	U	50.0	P	10/15/2024	14:39	LB132951
	Arsenic	20.0	+/-20.0	U	20.0	P	10/15/2024	14:39	LB132951
	Barium	100	+/-100	U	100	P	10/15/2024	14:39	LB132951
	Beryllium	6.00	+/-6.00	U	6.00	P	10/15/2024	14:39	LB132951
	Cadmium	6.00	+/-6.00	U	6.00	P	10/15/2024	14:39	LB132951
	Calcium	2000	+/-2000	U	2000	P	10/15/2024	14:39	LB132951
	Chromium	10.0	+/-10.0	U	10.0	P	10/15/2024	14:39	LB132951
	Cobalt	30.0	+/-30.0	U	30.0	P	10/15/2024	14:39	LB132951
	Copper	20.0	+/-20.0	U	20.0	P	10/15/2024	14:39	LB132951
	Iron	100	+/-100	U	100	P	10/15/2024	14:39	LB132951
	Lead	12.0	+/-12.0	U	12.0	P	10/15/2024	14:39	LB132951

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Roman E&G Corp		SDG No.: P4258							
Contract: ROMA02	Lab Code: CHEM	Case No.: P4258	SAS No.: P4258						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	10/15/2024	14:39	LB132951
	Manganese	20.0	+/-20.0	U	20.0	P	10/15/2024	14:39	LB132951
	Nickel	40.0	+/-40.0	U	40.0	P	10/15/2024	14:39	LB132951
	Potassium	2000	+/-2000	U	2000	P	10/15/2024	14:39	LB132951
	Selenium	20.0	+/-20.0	U	20.0	P	10/15/2024	14:39	LB132951
	Silver	10.0	+/-10.0	U	10.0	P	10/15/2024	14:39	LB132951
	Sodium	2000	+/-2000	U	2000	P	10/15/2024	14:39	LB132951
	Thallium	40.0	+/-40.0	U	40.0	P	10/15/2024	14:39	LB132951
	Vanadium	40.0	+/-40.0	U	40.0	P	10/15/2024	14:39	LB132951
	Zinc	40.0	+/-40.0	U	40.0	P	10/15/2024	14:39	LB132951
CCB05	Aluminum	100	+/-100	U	100	P	10/15/2024	15:30	LB132951
	Antimony	50.0	+/-50.0	U	50.0	P	10/15/2024	15:30	LB132951
	Arsenic	20.0	+/-20.0	U	20.0	P	10/15/2024	15:30	LB132951
	Barium	100	+/-100	U	100	P	10/15/2024	15:30	LB132951
	Beryllium	6.00	+/-6.00	U	6.00	P	10/15/2024	15:30	LB132951
	Cadmium	6.00	+/-6.00	U	6.00	P	10/15/2024	15:30	LB132951
	Calcium	2000	+/-2000	U	2000	P	10/15/2024	15:30	LB132951
	Chromium	10.0	+/-10.0	U	10.0	P	10/15/2024	15:30	LB132951
	Cobalt	30.0	+/-30.0	U	30.0	P	10/15/2024	15:30	LB132951
	Copper	20.0	+/-20.0	U	20.0	P	10/15/2024	15:30	LB132951
	Iron	100	+/-100	U	100	P	10/15/2024	15:30	LB132951
	Lead	12.0	+/-12.0	U	12.0	P	10/15/2024	15:30	LB132951
	Magnesium	2000	+/-2000	U	2000	P	10/15/2024	15:30	LB132951
	Manganese	20.0	+/-20.0	U	20.0	P	10/15/2024	15:30	LB132951
	Nickel	40.0	+/-40.0	U	40.0	P	10/15/2024	15:30	LB132951
	Potassium	2000	+/-2000	U	2000	P	10/15/2024	15:30	LB132951
	Selenium	20.0	+/-20.0	U	20.0	P	10/15/2024	15:30	LB132951
	Silver	10.0	+/-10.0	U	10.0	P	10/15/2024	15:30	LB132951
	Sodium	2000	+/-2000	U	2000	P	10/15/2024	15:30	LB132951
	Thallium	40.0	+/-40.0	U	40.0	P	10/15/2024	15:30	LB132951
	Vanadium	40.0	+/-40.0	U	40.0	P	10/15/2024	15:30	LB132951
	Zinc	40.0	+/-40.0	U	40.0	P	10/15/2024	15:30	LB132951
CCB06	Aluminum	100	+/-100	U	100	P	10/15/2024	16:21	LB132951
	Antimony	50.0	+/-50.0	U	50.0	P	10/15/2024	16:21	LB132951
	Arsenic	20.0	+/-20.0	U	20.0	P	10/15/2024	16:21	LB132951
	Barium	100	+/-100	U	100	P	10/15/2024	16:21	LB132951
	Beryllium	6.00	+/-6.00	U	6.00	P	10/15/2024	16:21	LB132951
	Cadmium	6.00	+/-6.00	U	6.00	P	10/15/2024	16:21	LB132951
	Calcium	2000	+/-2000	U	2000	P	10/15/2024	16:21	LB132951

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Roman E&G Corp		SDG No.: P4258							
Contract: ROMA02	Lab Code: CHEM	Case No.: P4258	SAS No.: P4258						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	10/15/2024	16:21	LB132951
	Cobalt	30.0	+/-30.0	U	30.0	P	10/15/2024	16:21	LB132951
	Copper	20.0	+/-20.0	U	20.0	P	10/15/2024	16:21	LB132951
	Iron	100	+/-100	U	100	P	10/15/2024	16:21	LB132951
	Lead	12.0	+/-12.0	U	12.0	P	10/15/2024	16:21	LB132951
	Magnesium	2000	+/-2000	U	2000	P	10/15/2024	16:21	LB132951
	Manganese	20.0	+/-20.0	U	20.0	P	10/15/2024	16:21	LB132951
	Nickel	40.0	+/-40.0	U	40.0	P	10/15/2024	16:21	LB132951
	Potassium	2000	+/-2000	U	2000	P	10/15/2024	16:21	LB132951
	Selenium	20.0	+/-20.0	U	20.0	P	10/15/2024	16:21	LB132951
	Silver	10.0	+/-10.0	U	10.0	P	10/15/2024	16:21	LB132951
	Sodium	2000	+/-2000	U	2000	P	10/15/2024	16:21	LB132951
	Thallium	40.0	+/-40.0	U	40.0	P	10/15/2024	16:21	LB132951
	Vanadium	40.0	+/-40.0	U	40.0	P	10/15/2024	16:21	LB132951
	Zinc	40.0	+/-40.0	U	40.0	P	10/15/2024	16:21	LB132951
CCB07	Aluminum	100	+/-100	U	100	P	10/15/2024	17:18	LB132951
	Antimony	50.0	+/-50.0	U	50.0	P	10/15/2024	17:18	LB132951
	Arsenic	20.0	+/-20.0	U	20.0	P	10/15/2024	17:18	LB132951
	Barium	100	+/-100	U	100	P	10/15/2024	17:18	LB132951
	Beryllium	6.00	+/-6.00	U	6.00	P	10/15/2024	17:18	LB132951
	Cadmium	6.00	+/-6.00	U	6.00	P	10/15/2024	17:18	LB132951
	Calcium	2000	+/-2000	U	2000	P	10/15/2024	17:18	LB132951
	Chromium	10.0	+/-10.0	U	10.0	P	10/15/2024	17:18	LB132951
	Cobalt	30.0	+/-30.0	U	30.0	P	10/15/2024	17:18	LB132951
	Copper	20.0	+/-20.0	U	20.0	P	10/15/2024	17:18	LB132951
	Iron	100	+/-100	U	100	P	10/15/2024	17:18	LB132951
	Lead	12.0	+/-12.0	U	12.0	P	10/15/2024	17:18	LB132951
	Magnesium	2000	+/-2000	U	2000	P	10/15/2024	17:18	LB132951
	Manganese	20.0	+/-20.0	U	20.0	P	10/15/2024	17:18	LB132951
	Nickel	40.0	+/-40.0	U	40.0	P	10/15/2024	17:18	LB132951
	Potassium	2000	+/-2000	U	2000	P	10/15/2024	17:18	LB132951
	Selenium	20.0	+/-20.0	U	20.0	P	10/15/2024	17:18	LB132951
	Silver	10.0	+/-10.0	U	10.0	P	10/15/2024	17:18	LB132951
	Sodium	2000	+/-2000	U	2000	P	10/15/2024	17:18	LB132951
	Thallium	40.0	+/-40.0	U	40.0	P	10/15/2024	17:18	LB132951
	Vanadium	40.0	+/-40.0	U	40.0	P	10/15/2024	17:18	LB132951
	Zinc	40.0	+/-40.0	U	40.0	P	10/15/2024	17:18	LB132951
CCB08	Aluminum	100	+/-100	U	100	P	10/15/2024	17:53	LB132951
	Antimony	50.0	+/-50.0	U	50.0	P	10/15/2024	17:53	LB132951

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Roman E&G Corp</u>		SDG No.: <u>P4258</u>							
Contract: <u>ROMA02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4258</u>	SAS No.: <u>P4258</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	10/15/2024	17:53	LB132951
	Barium	100	+/-100	U	100	P	10/15/2024	17:53	LB132951
	Beryllium	6.00	+/-6.00	U	6.00	P	10/15/2024	17:53	LB132951
	Cadmium	6.00	+/-6.00	U	6.00	P	10/15/2024	17:53	LB132951
	Calcium	2000	+/-2000	U	2000	P	10/15/2024	17:53	LB132951
	Chromium	10.0	+/-10.0	U	10.0	P	10/15/2024	17:53	LB132951
	Cobalt	30.0	+/-30.0	U	30.0	P	10/15/2024	17:53	LB132951
	Copper	20.0	+/-20.0	U	20.0	P	10/15/2024	17:53	LB132951
	Iron	100	+/-100	U	100	P	10/15/2024	17:53	LB132951
	Lead	12.0	+/-12.0	U	12.0	P	10/15/2024	17:53	LB132951
	Magnesium	2000	+/-2000	U	2000	P	10/15/2024	17:53	LB132951
	Manganese	20.0	+/-20.0	U	20.0	P	10/15/2024	17:53	LB132951
	Nickel	40.0	+/-40.0	U	40.0	P	10/15/2024	17:53	LB132951
	Potassium	2000	+/-2000	U	2000	P	10/15/2024	17:53	LB132951
	Selenium	20.0	+/-20.0	U	20.0	P	10/15/2024	17:53	LB132951
	Silver	10.0	+/-10.0	U	10.0	P	10/15/2024	17:53	LB132951
	Sodium	2000	+/-2000	U	2000	P	10/15/2024	17:53	LB132951
	Thallium	40.0	+/-40.0	U	40.0	P	10/15/2024	17:53	LB132951
	Vanadium	40.0	+/-40.0	U	40.0	P	10/15/2024	17:53	LB132951
	Zinc	40.0	+/-40.0	U	40.0	P	10/15/2024	17:53	LB132951

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Roman E&G Corp</u>		SDG No.: <u>P4258</u>							
Contract: <u>ROMA02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4258</u>	SAS No.: <u>P4258</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	10/18/2024	13:23	LB133014
	Antimony	50.0	+/-50.0	U	50.0	P	10/18/2024	13:23	LB133014
	Arsenic	20.0	+/-20.0	U	20.0	P	10/18/2024	13:23	LB133014
	Barium	100	+/-100	U	100	P	10/18/2024	13:23	LB133014
	Beryllium	6.00	+/-6.00	U	6.00	P	10/18/2024	13:23	LB133014
	Cadmium	6.00	+/-6.00	U	6.00	P	10/18/2024	13:23	LB133014
	Calcium	2000	+/-2000	U	2000	P	10/18/2024	13:23	LB133014
	Chromium	10.0	+/-10.0	U	10.0	P	10/18/2024	13:23	LB133014
	Cobalt	30.0	+/-30.0	U	30.0	P	10/18/2024	13:23	LB133014
	Copper	20.0	+/-20.0	U	20.0	P	10/18/2024	13:23	LB133014
	Iron	100	+/-100	U	100	P	10/18/2024	13:23	LB133014
	Lead	12.0	+/-12.0	U	12.0	P	10/18/2024	13:23	LB133014
	Magnesium	2000	+/-2000	U	2000	P	10/18/2024	13:23	LB133014
	Manganese	20.0	+/-20.0	U	20.0	P	10/18/2024	13:23	LB133014
	Nickel	40.0	+/-40.0	U	40.0	P	10/18/2024	13:23	LB133014
	Potassium	2000	+/-2000	U	2000	P	10/18/2024	13:23	LB133014
	Selenium	20.0	+/-20.0	U	20.0	P	10/18/2024	13:23	LB133014
	Silver	10.0	+/-10.0	U	10.0	P	10/18/2024	13:23	LB133014
	Sodium	2000	+/-2000	U	2000	P	10/18/2024	13:23	LB133014
	Thallium	40.0	+/-40.0	U	40.0	P	10/18/2024	13:23	LB133014
	Vanadium	40.0	+/-40.0	U	40.0	P	10/18/2024	13:23	LB133014
	Zinc	40.0	+/-40.0	U	40.0	P	10/18/2024	13:23	LB133014

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Roman E&G Corp		SDG No.: P4258							
Contract: ROMA02	Lab Code: CHEM	Case No.: P4258	SAS No.: P4258						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	10/18/2024	14:05	LB133014
	Antimony	50.0	+/-50.0	U	50.0	P	10/18/2024	14:05	LB133014
	Arsenic	20.0	+/-20.0	U	20.0	P	10/18/2024	14:05	LB133014
	Barium	100	+/-100	U	100	P	10/18/2024	14:05	LB133014
	Beryllium	6.00	+/-6.00	U	6.00	P	10/18/2024	14:05	LB133014
	Cadmium	0.48	+/-6.00	J	6.00	P	10/18/2024	14:05	LB133014
	Calcium	2000	+/-2000	U	2000	P	10/18/2024	14:05	LB133014
	Chromium	10.0	+/-10.0	U	10.0	P	10/18/2024	14:05	LB133014
	Cobalt	30.0	+/-30.0	U	30.0	P	10/18/2024	14:05	LB133014
	Copper	20.0	+/-20.0	U	20.0	P	10/18/2024	14:05	LB133014
	Iron	100	+/-100	U	100	P	10/18/2024	14:05	LB133014
	Lead	12.0	+/-12.0	U	12.0	P	10/18/2024	14:05	LB133014
	Magnesium	2000	+/-2000	U	2000	P	10/18/2024	14:05	LB133014
	Manganese	20.0	+/-20.0	U	20.0	P	10/18/2024	14:05	LB133014
	Nickel	40.0	+/-40.0	U	40.0	P	10/18/2024	14:05	LB133014
	Potassium	2000	+/-2000	U	2000	P	10/18/2024	14:05	LB133014
	Selenium	20.0	+/-20.0	U	20.0	P	10/18/2024	14:05	LB133014
	Silver	10.0	+/-10.0	U	10.0	P	10/18/2024	14:05	LB133014
	Sodium	2000	+/-2000	U	2000	P	10/18/2024	14:05	LB133014
	Thallium	40.0	+/-40.0	U	40.0	P	10/18/2024	14:05	LB133014
	Vanadium	40.0	+/-40.0	U	40.0	P	10/18/2024	14:05	LB133014
	Zinc	40.0	+/-40.0	U	40.0	P	10/18/2024	14:05	LB133014
CCB02	Aluminum	100	+/-100	U	100	P	10/18/2024	14:57	LB133014
	Antimony	50.0	+/-50.0	U	50.0	P	10/18/2024	14:57	LB133014
	Arsenic	20.0	+/-20.0	U	20.0	P	10/18/2024	14:57	LB133014
	Barium	100	+/-100	U	100	P	10/18/2024	14:57	LB133014
	Beryllium	6.00	+/-6.00	U	6.00	P	10/18/2024	14:57	LB133014
	Cadmium	6.00	+/-6.00	U	6.00	P	10/18/2024	14:57	LB133014
	Calcium	2000	+/-2000	U	2000	P	10/18/2024	14:57	LB133014
	Chromium	10.0	+/-10.0	U	10.0	P	10/18/2024	14:57	LB133014
	Cobalt	30.0	+/-30.0	U	30.0	P	10/18/2024	14:57	LB133014
	Copper	20.0	+/-20.0	U	20.0	P	10/18/2024	14:57	LB133014
	Iron	100	+/-100	U	100	P	10/18/2024	14:57	LB133014
	Lead	12.0	+/-12.0	U	12.0	P	10/18/2024	14:57	LB133014
	Magnesium	2000	+/-2000	U	2000	P	10/18/2024	14:57	LB133014
	Manganese	20.0	+/-20.0	U	20.0	P	10/18/2024	14:57	LB133014
	Nickel	40.0	+/-40.0	U	40.0	P	10/18/2024	14:57	LB133014
	Potassium	2000	+/-2000	U	2000	P	10/18/2024	14:57	LB133014
	Selenium	20.0	+/-20.0	U	20.0	P	10/18/2024	14:57	LB133014

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Roman E&G Corp		SDG No.: P4258							
Contract: ROMA02	Lab Code: CHEM	Case No.: P4258	SAS No.: P4258						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	10/18/2024	14:57	LB133014
	Sodium	2000	+/-2000	U	2000	P	10/18/2024	14:57	LB133014
	Thallium	40.0	+/-40.0	U	40.0	P	10/18/2024	14:57	LB133014
	Vanadium	40.0	+/-40.0	U	40.0	P	10/18/2024	14:57	LB133014
	Zinc	40.0	+/-40.0	U	40.0	P	10/18/2024	14:57	LB133014
CCB03	Aluminum	100	+/-100	U	100	P	10/18/2024	16:39	LB133014
	Antimony	50.0	+/-50.0	U	50.0	P	10/18/2024	16:39	LB133014
	Arsenic	20.0	+/-20.0	U	20.0	P	10/18/2024	16:39	LB133014
	Barium	100	+/-100	U	100	P	10/18/2024	16:39	LB133014
	Beryllium	6.00	+/-6.00	U	6.00	P	10/18/2024	16:39	LB133014
	Cadmium	6.00	+/-6.00	U	6.00	P	10/18/2024	16:39	LB133014
	Calcium	2000	+/-2000	U	2000	P	10/18/2024	16:39	LB133014
	Chromium	10.0	+/-10.0	U	10.0	P	10/18/2024	16:39	LB133014
	Cobalt	30.0	+/-30.0	U	30.0	P	10/18/2024	16:39	LB133014
	Copper	20.0	+/-20.0	U	20.0	P	10/18/2024	16:39	LB133014
	Iron	100	+/-100	U	100	P	10/18/2024	16:39	LB133014
	Lead	12.0	+/-12.0	U	12.0	P	10/18/2024	16:39	LB133014
	Magnesium	2000	+/-2000	U	2000	P	10/18/2024	16:39	LB133014
	Manganese	20.0	+/-20.0	U	20.0	P	10/18/2024	16:39	LB133014
	Nickel	40.0	+/-40.0	U	40.0	P	10/18/2024	16:39	LB133014
	Potassium	2000	+/-2000	U	2000	P	10/18/2024	16:39	LB133014
	Selenium	20.0	+/-20.0	U	20.0	P	10/18/2024	16:39	LB133014
	Silver	10.0	+/-10.0	U	10.0	P	10/18/2024	16:39	LB133014
	Sodium	2000	+/-2000	U	2000	P	10/18/2024	16:39	LB133014
	Thallium	40.0	+/-40.0	U	40.0	P	10/18/2024	16:39	LB133014
	Vanadium	40.0	+/-40.0	U	40.0	P	10/18/2024	16:39	LB133014
	Zinc	40.0	+/-40.0	U	40.0	P	10/18/2024	16:39	LB133014
CCB04	Aluminum	100	+/-100	U	100	P	10/18/2024	17:32	LB133014
	Antimony	50.0	+/-50.0	U	50.0	P	10/18/2024	17:32	LB133014
	Arsenic	20.0	+/-20.0	U	20.0	P	10/18/2024	17:32	LB133014
	Barium	100	+/-100	U	100	P	10/18/2024	17:32	LB133014
	Beryllium	6.00	+/-6.00	U	6.00	P	10/18/2024	17:32	LB133014
	Cadmium	6.00	+/-6.00	U	6.00	P	10/18/2024	17:32	LB133014
	Calcium	2000	+/-2000	U	2000	P	10/18/2024	17:32	LB133014
	Chromium	10.0	+/-10.0	U	10.0	P	10/18/2024	17:32	LB133014
	Cobalt	30.0	+/-30.0	U	30.0	P	10/18/2024	17:32	LB133014
	Copper	20.0	+/-20.0	U	20.0	P	10/18/2024	17:32	LB133014
	Iron	100	+/-100	U	100	P	10/18/2024	17:32	LB133014
	Lead	12.0	+/-12.0	U	12.0	P	10/18/2024	17:32	LB133014

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Roman E&G Corp **SDG No.:** P4258
Contract: ROMA02 **Lab Code:** CHEM **Case No.:** P4258 **SAS No.:** P4258

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	10/18/2024	17:32	LB133014
	Manganese	20.0	+/-20.0	U	20.0	P	10/18/2024	17:32	LB133014
	Nickel	40.0	+/-40.0	U	40.0	P	10/18/2024	17:32	LB133014
	Potassium	2000	+/-2000	U	2000	P	10/18/2024	17:32	LB133014
	Selenium	20.0	+/-20.0	U	20.0	P	10/18/2024	17:32	LB133014
	Silver	10.0	+/-10.0	U	10.0	P	10/18/2024	17:32	LB133014
	Sodium	2000	+/-2000	U	2000	P	10/18/2024	17:32	LB133014
	Thallium	40.0	+/-40.0	U	40.0	P	10/18/2024	17:32	LB133014
	Vanadium	40.0	+/-40.0	U	40.0	P	10/18/2024	17:32	LB133014
	Zinc	40.0	+/-40.0	U	40.0	P	10/18/2024	17:32	LB133014