

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Roman E&G Corp

SDG No.: P4258

Contract: ROMA02

Lab Code: CHEM

Case No.: P4258

SAS No.: P4258

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB163879BS Mercury	ug/L	4.0	4.07		102	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Roman E&G Corp **SDG No.:** P4258
Contract: ROMA02 **Lab Code:** CHEM **Case No.:** P4258 **SAS No.:** P4258

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB163885BS							
Arsenic	ug/L	4000	4020		100	80 - 120	P
Barium	ug/L	1000	1080		108	80 - 120	P
Cadmium	ug/L	1000	986		99	80 - 120	P
Chromium	ug/L	2000	2100		105	80 - 120	P
Copper	ug/L	1500	1680		112	80 - 120	P
Lead	ug/L	5000	5000		100	80 - 120	P
Nickel	ug/L	2500	2670		107	80 - 120	P
Selenium	ug/L	10000	10200		102	80 - 120	P
Silver	ug/L	380	405		107	80 - 120	P
Zinc	ug/L	1000	1110		111	80 - 120	P