

Continuing Calibration Report for SequenceID : FC101124AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FC067420.D

|                   |              |         |        |        |            |            |       |
|-------------------|--------------|---------|--------|--------|------------|------------|-------|
| Aliphatic C9-C12  | 7672816.000  | 60.000  | 3.175  | 6.457  | 127880.267 | 128354.488 | 0.369 |
| Aliphatic C12-C16 | 5034510.000  | 40.000  | 6.458  | 9.848  | 125862.750 | 130957.677 | 3.891 |
| Aliphatic C16-C21 | 7291288.000  | 60.000  | 9.849  | 13.207 | 121521.467 | 131476.102 | 7.571 |
| Aliphatic C21-C28 | 9260604.000  | 80.000  | 13.208 | 16.862 | 115757.550 | 126823.664 | 8.726 |
| Aliphatic C28-C40 | 11796013.000 | 120.000 | 16.863 | 21.712 | 98300.108  | 105215.848 | 6.573 |
| Aliphatic EPH     | 41055231.000 | 360.000 | 3.175  | 21.712 | 114042.308 | 121110.937 | 5.836 |