

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Roman E&G Corp
Lab Code: CHEM **Cas No:** P4258 **SAS No :** P4258 **SDG No:** P4258
Sample No : PB164054BS **Datafile:** FC067410.D
Client ID : PB164054BS

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aliphatic C9-C12	10.0	5.95	60		(40-140)
Aliphatic C12-C16	6.7	5.05	76		(40-140)
Aliphatic C16-C21	10.0	7.93	79		(40-140)
Aliphatic C21-C28	13.3	10.6	80		(40-140)
Aliphatic C28-C40	20.0	15.4	77		(40-140)