

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

Continuing Calib Date: 10/11/2024 Initial Calibration Date(s): 10/07/2024 10/07/2024

Continuing Calib Time: 10:37 Initial Calibration Time(s): 11:12 12:47

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.27	7.28	7.18	7.38	0.01
2,4-DCAA	7.10	7.10	7.00	7.20	0.01
DICHLORPROP	7.96	7.96	7.86	8.06	0.00
2,4-D	8.18	8.19	8.09	8.29	0.01
2,4,5-TP(Silvex)	9.03	9.04	8.94	9.14	0.01
2,4,5-T	9.32	9.32	9.22	9.42	0.00
2,4-DB	9.88	9.89	9.79	9.99	0.01
Dinoseb	11.06	11.06	10.96	11.16	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

Continuing Calib Date: 10/11/2024 Initial Calibration Date(s): 10/07/2024 10/07/2024

Continuing Calib Time: 10:37 Initial Calibration Time(s): 11:12 12:47

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.82	7.82	7.72	7.92	0.00
2,4-DCAA	7.62	7.63	7.53	7.73	0.01
DICHLORPROP	8.52	8.53	8.43	8.63	0.01
2,4-D	8.85	8.85	8.75	8.95	0.00
2,4,5-TP(Silvex)	9.74	9.74	9.64	9.84	0.00
2,4,5-T	10.15	10.16	10.06	10.26	0.01
2,4-DB	10.71	10.72	10.62	10.82	0.01
Dinoseb	11.09	11.09	10.99	11.19	0.00