

### CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 10/07/2024 10/07/2024

Client Sample No.: CCAL01 Date Analyzed: 10/11/2024

Lab Sample No.: HSTDCCC750 Data File : PS027905.D Time Analyzed: 10:37

| COMPOUND         | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|------------------|--------|-----------|--------|--------------------|-------------------|------|
|                  |        | FROM      | TO     |                    |                   |      |
| 2,4,5-T          | 9.318  | 9.223     | 9.423  | 719.450            | 712.500           | 1.0  |
| 2,4,5-TP(Silvex) | 9.034  | 8.938     | 9.138  | 717.440            | 712.500           | 0.7  |
| 2,4-D            | 8.181  | 8.085     | 8.285  | 686.600            | 705.000           | -2.6 |
| 2,4-DB           | 9.881  | 9.785     | 9.985  | 702.270            | 712.500           | -1.4 |
| 2,4-DCAA         | 7.095  | 6.998     | 7.198  | 735.100            | 750.000           | -2.0 |
| DICAMBA          | 7.274  | 7.177     | 7.377  | 705.590            | 705.000           | 0.1  |
| DICHLORPROP      | 7.959  | 7.863     | 8.063  | 695.860            | 705.000           | -1.3 |
| Dinoseb          | 11.057 | 10.961    | 11.161 | 707.700            | 705.000           | 0.4  |

### CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/07/2024 10/07/2024

Client Sample No.: CCAL01 Date Analyzed: 10/11/2024

Lab Sample No.: HSTDCCC750 Data File : PS027905.D Time Analyzed: 10:37

| COMPOUND         | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|------------------|--------|-----------|--------|--------------------|-------------------|------|
|                  |        | FROM      | TO     |                    |                   |      |
| 2,4,5-T          | 10.153 | 10.058    | 10.258 | 708.250            | 712.500           | -0.6 |
| 2,4,5-TP(Silvex) | 9.739  | 9.643     | 9.843  | 721.720            | 712.500           | 1.3  |
| 2,4-D            | 8.849  | 8.752     | 8.952  | 672.380            | 705.000           | -4.6 |
| 2,4-DB           | 10.714 | 10.619    | 10.819 | 705.100            | 712.500           | -1.0 |
| 2,4-DCAA         | 7.622  | 7.528     | 7.728  | 698.610            | 750.000           | -6.9 |
| DICAMBA          | 7.817  | 7.722     | 7.922  | 732.660            | 705.000           | 3.9  |
| DICHLORPROP      | 8.524  | 8.428     | 8.628  | 690.830            | 705.000           | -2.0 |
| Dinoseb          | 11.088 | 10.993    | 11.193 | 705.540            | 705.000           | 0.1  |