

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 10/07/2024 10/07/2024

Client Sample No.: CCAL02 Date Analyzed: 10/11/2024

Lab Sample No.: HSTDCCC750 Data File : PS027916.D Time Analyzed: 21:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.319	9.223	9.423	723.470	712.500	1.5
2,4,5-TP(Silvex)	9.035	8.938	9.138	715.280	712.500	0.4
2,4-D	8.182	8.085	8.285	691.690	705.000	-1.9
2,4-DB	9.881	9.785	9.985	706.900	712.500	-0.8
2,4-DCAA	7.095	6.998	7.198	738.190	750.000	-1.6
DICAMBA	7.274	7.177	7.377	705.560	705.000	0.1
DICHLORPROP	7.959	7.863	8.063	697.370	705.000	-1.1
Dinoseb	11.057	10.961	11.161	717.240	705.000	1.7

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/07/2024 10/07/2024

Client Sample No.: CCAL02 Date Analyzed: 10/11/2024

Lab Sample No.: HSTDCCC750 Data File : PS027916.D Time Analyzed: 21:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.153	10.058	10.258	720.170	712.500	1.1
2,4,5-TP(Silvex)	9.739	9.643	9.843	701.080	712.500	-1.6
2,4-D	8.848	8.752	8.952	705.930	705.000	0.1
2,4-DB	10.713	10.619	10.819	697.420	712.500	-2.1
2,4-DCAA	7.623	7.528	7.728	675.590	750.000	-9.9
DICAMBA	7.817	7.722	7.922	708.800	705.000	0.5
DICHLORPROP	8.524	8.428	8.628	679.310	705.000	-3.6
Dinoseb	11.089	10.993	11.193	702.130	705.000	-0.4