



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4258

Client: Roman E&G Corp

Analytical Method: 8151A Datafile : PS027907.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB164060BS	DICAMBA	166.6	160	ug/Kg	96				72	129	
	DICHLORPROP	166.6	170	ug/Kg	102				77	135	
	2,4-D	166.6	197	ug/Kg	118				65	144	
	2,4,5-TP(Silvex)	166.6	176	ug/Kg	106				74	146	
	2,4,5-T	166.6	164	ug/Kg	98				77	134	
	2,4-DB	166.6	186	ug/Kg	112				72	122	
	Dinoseb	166.6	162	ug/Kg	97				74	132	