

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

Continuing Calib Date: 10/02/2024 Initial Calibration Date(s): 09/26/2024 09/26/2024

Continuing Calib Time: 09:04 Initial Calibration Time(s): 16:35 23:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.92	5.92	5.82	6.02	0.00
Aroclor-1016-2	(2)	5.94	5.94	5.84	6.04	0.00
Aroclor-1016-3	(3)	6.01	6.00	5.90	6.10	-0.01
Aroclor-1016-4	(4)	6.10	6.10	6.00	6.20	0.00
Aroclor-1016-5	(5)	6.40	6.40	6.30	6.50	0.00
Aroclor-1260-1	(1)	7.52	7.52	7.42	7.62	0.00
Aroclor-1260-2	(2)	7.77	7.77	7.67	7.87	0.00
Aroclor-1260-3	(3)	8.14	8.13	8.03	8.23	-0.01
Aroclor-1260-4	(4)	8.37	8.37	8.27	8.47	0.00
Aroclor-1260-5	(5)	8.71	8.71	8.61	8.81	0.00
Tetrachloro-m-xylene		4.76	4.75	4.65	4.85	-0.01
Decachlorobiphenyl		10.67	10.66	10.56	10.76	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

Continuing Calib Date: 10/02/2024 Initial Calibration Date(s): 09/26/2024 09/26/2024

Continuing Calib Time: 09:04 Initial Calibration Time(s): 16:35 23:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.16	5.16	5.06	5.26	0.00
Aroclor-1016-2	(2)	5.18	5.18	5.08	5.28	0.00
Aroclor-1016-3	(3)	5.36	5.37	5.27	5.47	0.01
Aroclor-1016-4	(4)	5.40	5.40	5.30	5.50	0.00
Aroclor-1016-5	(5)	5.62	5.62	5.52	5.72	0.00
Aroclor-1260-1	(1)	6.67	6.67	6.57	6.77	0.00
Aroclor-1260-2	(2)	6.85	6.86	6.76	6.96	0.01
Aroclor-1260-3	(3)	7.01	7.01	6.91	7.11	0.00
Aroclor-1260-4	(4)	7.49	7.49	7.39	7.59	0.00
Aroclor-1260-5	(5)	7.72	7.73	7.63	7.83	0.01
Tetrachloro-m-xylene		4.05	4.05	3.95	4.15	0.00
Decachlorobiphenyl		9.22	9.23	9.13	9.33	0.01