

### CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

Continuing Calib Date: 10/02/2024 Initial Calibration Date(s): 09/26/2024 09/26/2024

Continuing Calib Time: 13:34 Initial Calibration Time(s): 16:35 23:51

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1016-1       | (1) | 5.92       | 5.92      | 5.82      | 6.02  | 0.00       |
| Aroclor-1016-2       | (2) | 5.94       | 5.94      | 5.84      | 6.04  | 0.00       |
| Aroclor-1016-3       | (3) | 6.00       | 6.00      | 5.90      | 6.10  | 0.00       |
| Aroclor-1016-4       | (4) | 6.10       | 6.10      | 6.00      | 6.20  | 0.00       |
| Aroclor-1016-5       | (5) | 6.40       | 6.40      | 6.30      | 6.50  | 0.00       |
| Aroclor-1260-1       | (1) | 7.52       | 7.52      | 7.42      | 7.62  | 0.00       |
| Aroclor-1260-2       | (2) | 7.77       | 7.77      | 7.67      | 7.87  | 0.00       |
| Aroclor-1260-3       | (3) | 8.13       | 8.13      | 8.03      | 8.23  | 0.00       |
| Aroclor-1260-4       | (4) | 8.37       | 8.37      | 8.27      | 8.47  | 0.00       |
| Aroclor-1260-5       | (5) | 8.71       | 8.71      | 8.61      | 8.81  | 0.00       |
| Tetrachloro-m-xylene |     | 4.75       | 4.75      | 4.65      | 4.85  | 0.00       |
| Decachlorobiphenyl   |     | 10.66      | 10.66     | 10.56     | 10.76 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

Continuing Calib Date: 10/02/2024 Initial Calibration Date(s): 09/26/2024 09/26/2024

Continuing Calib Time: 13:34 Initial Calibration Time(s): 16:35 23:51

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|------------------------|------|------------|
| Aroclor-1016-1       | (1) | 5.16       | 5.16      | 5.06                   | 5.26 | 0.00       |
| Aroclor-1016-2       | (2) | 5.18       | 5.18      | 5.08                   | 5.28 | 0.00       |
| Aroclor-1016-3       | (3) | 5.36       | 5.37      | 5.27                   | 5.47 | 0.01       |
| Aroclor-1016-4       | (4) | 5.40       | 5.40      | 5.30                   | 5.50 | 0.00       |
| Aroclor-1016-5       | (5) | 5.62       | 5.62      | 5.52                   | 5.72 | 0.00       |
| Aroclor-1260-1       | (1) | 6.66       | 6.67      | 6.57                   | 6.77 | 0.01       |
| Aroclor-1260-2       | (2) | 6.85       | 6.86      | 6.76                   | 6.96 | 0.01       |
| Aroclor-1260-3       | (3) | 7.01       | 7.01      | 6.91                   | 7.11 | 0.00       |
| Aroclor-1260-4       | (4) | 7.48       | 7.49      | 7.39                   | 7.59 | 0.01       |
| Aroclor-1260-5       | (5) | 7.72       | 7.73      | 7.63                   | 7.83 | 0.01       |
| Tetrachloro-m-xylene |     | 4.05       | 4.05      | 3.95                   | 4.15 | 0.00       |
| Decachlorobiphenyl   |     | 9.22       | 9.23      | 9.13                   | 9.33 | 0.01       |