

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM **Case No.:** P4258 **SAS No.:** P4258 **SDG NO.:** P4258

Continuing Calib Date: 10/02/2024 **Initial Calibration Date(s):** 09/26/2024 09/26/2024

Continuing Calib Time: 14:06 **Initial Calibration Time(s):** 16:35 23:51

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.92	5.92	5.82	6.02	0.00
Aroclor-1248-2 (2)	6.19	6.19	6.09	6.29	0.00
Aroclor-1248-3 (3)	6.40	6.40	6.30	6.50	0.01
Aroclor-1248-4 (4)	6.80	6.80	6.70	6.90	0.00
Aroclor-1248-5 (5)	6.83	6.84	6.74	6.94	0.01
Tetrachloro-m-xylene	4.75	4.75	4.65	4.85	0.00
Decachlorobiphenyl	10.66	10.66	10.56	10.76	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM **Case No.:** P4258 **SAS No.:** P4258 **SDG NO.:** P4258

Continuing Calib Date: 10/02/2024 **Initial Calibration Date(s):** 09/26/2024 09/26/2024

Continuing Calib Time: 14:06 **Initial Calibration Time(s):** 16:35 23:51

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.16	5.16	5.06	5.26	0.00
Aroclor-1248-2 (2)	5.40	5.40	5.30	5.50	0.00
Aroclor-1248-3 (3)	5.44	5.45	5.35	5.55	0.01
Aroclor-1248-4 (4)	5.62	5.62	5.52	5.72	0.00
Aroclor-1248-5 (5)	6.02	6.02	5.92	6.12	0.00
Tetrachloro-m-xylene	4.05	4.05	3.95	4.15	0.00
Decachlorobiphenyl	9.22	9.22	9.12	9.32	0.00