

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

Continuing Calib Date: 10/03/2024 Initial Calibration Date(s): 09/26/2024 09/26/2024

Continuing Calib Time: 02:54 Initial Calibration Time(s): 16:35 23:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	6.77	6.77	6.67	6.87	0.00
Aroclor-1254-2	(2)	6.99	6.99	6.89	7.09	0.00
Aroclor-1254-3	(3)	7.36	7.35	7.25	7.45	-0.01
Aroclor-1254-4	(4)	7.64	7.64	7.54	7.74	0.00
Aroclor-1254-5	(5)	8.06	8.05	7.95	8.15	-0.01
Tetrachloro-m-xylene		4.75	4.75	4.65	4.85	0.00
Decachlorobiphenyl		10.66	10.66	10.56	10.76	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

Continuing Calib Date: 10/03/2024 Initial Calibration Date(s): 09/26/2024 09/26/2024

Continuing Calib Time: 02:54 Initial Calibration Time(s): 16:35 23:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	5.98	5.98	5.88	6.08	0.01
Aroclor-1254-2	(2)	6.12	6.13	6.03	6.23	0.01
Aroclor-1254-3	(3)	6.53	6.54	6.44	6.64	0.01
Aroclor-1254-4	(4)	6.76	6.77	6.67	6.87	0.01
Aroclor-1254-5	(5)	7.18	7.19	7.09	7.29	0.01
Tetrachloro-m-xylene		4.05	4.05	3.95	4.15	0.00
Decachlorobiphenyl		9.22	9.23	9.13	9.33	0.01