

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/26/2024 09/26/2024

Client Sample No.: CCAL01 Date Analyzed: 10/02/2024

Lab Sample No.: AR1660CCC500 Data File : PP067426.D Time Analyzed: 13:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.917	5.817	6.017	474.830	500.000	-5.0
Aroclor-1016-2	5.940	5.840	6.040	469.370	500.000	-6.1
Aroclor-1016-3	6.003	5.903	6.103	478.840	500.000	-4.2
Aroclor-1016-4	6.101	6.002	6.202	478.430	500.000	-4.3
Aroclor-1016-5	6.396	6.296	6.496	480.050	500.000	-4.0
Aroclor-1260-1	7.519	7.418	7.618	430.280	500.000	-13.9
Aroclor-1260-2	7.772	7.672	7.872	420.850	500.000	-15.8
Aroclor-1260-3	8.133	8.033	8.233	431.910	500.000	-13.6
Aroclor-1260-4	8.372	8.271	8.471	443.950	500.000	-11.2
Aroclor-1260-5	8.708	8.608	8.808	444.910	500.000	-11.0
Decachlorobiphenyl	10.661	10.561	10.761	47.300	50.000	-5.4
Tetrachloro-m-xylene	4.753	4.653	4.853	50.950	50.000	1.9

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/26/2024 09/26/2024

Client Sample No.: CCAL01 Date Analyzed: 10/02/2024

Lab Sample No.: AR1660CCC500 Data File : PP067426.D Time Analyzed: 13:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.160	5.064	5.264	449.100	500.000	-10.2
Aroclor-1016-2	5.181	5.084	5.284	459.490	500.000	-8.1
Aroclor-1016-3	5.361	5.265	5.465	460.030	500.000	-8.0
Aroclor-1016-4	5.400	5.304	5.504	457.300	500.000	-8.5
Aroclor-1016-5	5.619	5.524	5.724	462.480	500.000	-7.5
Aroclor-1260-1	6.664	6.569	6.769	433.030	500.000	-13.4
Aroclor-1260-2	6.850	6.755	6.955	436.360	500.000	-12.7
Aroclor-1260-3	7.008	6.913	7.113	439.810	500.000	-12.0
Aroclor-1260-4	7.483	7.388	7.588	446.520	500.000	-10.7
Aroclor-1260-5	7.721	7.626	7.826	454.610	500.000	-9.1
Decachlorobiphenyl	9.218	9.126	9.326	48.380	50.000	-3.2
Tetrachloro-m-xylene	4.051	3.953	4.153	49.670	50.000	-0.7