

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/26/2024 09/26/2024

Client Sample No.: CCAL02 Date Analyzed: 10/02/2024

Lab Sample No.: AR1660CCC500 Data File : PP067441.D Time Analyzed: 20:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.915	5.817	6.017	467.930	500.000	-6.4
Aroclor-1016-2	5.939	5.840	6.040	469.650	500.000	-6.1
Aroclor-1016-3	6.002	5.903	6.103	472.860	500.000	-5.4
Aroclor-1016-4	6.099	6.002	6.202	475.230	500.000	-5.0
Aroclor-1016-5	6.394	6.296	6.496	471.350	500.000	-5.7
Aroclor-1260-1	7.517	7.418	7.618	403.120	500.000	-19.4
Aroclor-1260-2	7.771	7.672	7.872	388.800	500.000	-22.2
Aroclor-1260-3	8.132	8.033	8.233	392.700	500.000	-21.5
Aroclor-1260-4	8.369	8.271	8.471	401.410	500.000	-19.7
Aroclor-1260-5	8.706	8.608	8.808	414.930	500.000	-17.0
Decachlorobiphenyl	10.662	10.561	10.761	45.470	50.000	-9.1
Tetrachloro-m-xylene	4.752	4.653	4.853	51.050	50.000	2.1

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/26/2024 09/26/2024

Client Sample No.: CCAL02 Date Analyzed: 10/02/2024

Lab Sample No.: AR1660CCC500 Data File : PP067441.D Time Analyzed: 20:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.160	5.064	5.264	461.960	500.000	-7.6
Aroclor-1016-2	5.180	5.084	5.284	464.100	500.000	-7.2
Aroclor-1016-3	5.361	5.265	5.465	487.090	500.000	-2.6
Aroclor-1016-4	5.400	5.304	5.504	473.080	500.000	-5.4
Aroclor-1016-5	5.619	5.524	5.724	488.330	500.000	-2.3
Aroclor-1260-1	6.664	6.569	6.769	426.610	500.000	-14.7
Aroclor-1260-2	6.850	6.755	6.955	421.090	500.000	-15.8
Aroclor-1260-3	7.008	6.913	7.113	413.950	500.000	-17.2
Aroclor-1260-4	7.483	7.388	7.588	420.960	500.000	-15.8
Aroclor-1260-5	7.721	7.626	7.826	422.840	500.000	-15.4
Decachlorobiphenyl	9.217	9.126	9.326	46.090	50.000	-7.8
Tetrachloro-m-xylene	4.051	3.953	4.153	50.060	50.000	0.1