

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/26/2024 09/26/2024

Client Sample No.: CCAL03 Date Analyzed: 10/03/2024

Lab Sample No.: AR1660CCC500 Data File : PP067452.D Time Analyzed: 02:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.916	5.817	6.017	471.040	500.000	-5.8
Aroclor-1016-2	5.938	5.840	6.040	477.400	500.000	-4.5
Aroclor-1016-3	6.002	5.903	6.103	481.710	500.000	-3.7
Aroclor-1016-4	6.100	6.002	6.202	473.420	500.000	-5.3
Aroclor-1016-5	6.395	6.296	6.496	479.060	500.000	-4.2
Aroclor-1260-1	7.518	7.418	7.618	439.290	500.000	-12.1
Aroclor-1260-2	7.771	7.672	7.872	419.560	500.000	-16.1
Aroclor-1260-3	8.132	8.033	8.233	424.750	500.000	-15.1
Aroclor-1260-4	8.370	8.271	8.471	433.350	500.000	-13.3
Aroclor-1260-5	8.707	8.608	8.808	438.610	500.000	-12.3
Decachlorobiphenyl	10.661	10.561	10.761	48.020	50.000	-4.0
Tetrachloro-m-xylene	4.752	4.653	4.853	51.400	50.000	2.8

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/26/2024 09/26/2024

Client Sample No.: CCAL03 Date Analyzed: 10/03/2024

Lab Sample No.: AR1660CCC500 Data File : PP067452.D Time Analyzed: 02:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.160	5.064	5.264	469.300	500.000	-6.1
Aroclor-1016-2	5.180	5.084	5.284	523.630	500.000	4.7
Aroclor-1016-3	5.361	5.265	5.465	513.240	500.000	2.6
Aroclor-1016-4	5.400	5.304	5.504	494.490	500.000	-1.1
Aroclor-1016-5	5.619	5.524	5.724	514.710	500.000	2.9
Aroclor-1260-1	6.663	6.569	6.769	449.540	500.000	-10.1
Aroclor-1260-2	6.849	6.755	6.955	450.550	500.000	-9.9
Aroclor-1260-3	7.007	6.913	7.113	445.980	500.000	-10.8
Aroclor-1260-4	7.482	7.388	7.588	458.770	500.000	-8.2
Aroclor-1260-5	7.720	7.626	7.826	463.040	500.000	-7.4
Decachlorobiphenyl	9.217	9.126	9.326	49.610	50.000	-0.8
Tetrachloro-m-xylene	4.051	3.953	4.153	50.720	50.000	1.4