

Surrogate Summary

SDG No.: **P4258**

Client: **Roman E&G Corp**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP067251.D	PIBLK-PP067251.D	Tetrachloro-m-xylene	1	20	18.8	94		60	140
		Decachlorobiphenyl	1	20	21.3	106		60	140
		Tetrachloro-m-xylene	2	20	18.6	93		60	140
		Decachlorobiphenyl	2	20	21.2	106		60	140
I.BLK-PP067430.D	PIBLK-PP067430.D	Tetrachloro-m-xylene	1	20	16.2	81		60	140
		Decachlorobiphenyl	1	20	16.4	82		60	140
		Tetrachloro-m-xylene	2	20	16.0	80		60	140
		Decachlorobiphenyl	2	20	16.7	83		60	140
PB163832BL	PB163832BL	Tetrachloro-m-xylene	1	20	16.1	81		32	144
		Decachlorobiphenyl	1	20	16.8	84		32	175
		Tetrachloro-m-xylene	2	20	15.9	80		32	144
		Decachlorobiphenyl	2	20	16.3	81		32	175
PB163832BS	PB163832BS	Tetrachloro-m-xylene	1	20	17.4	87		32	144
		Decachlorobiphenyl	1	20	16.1	81		32	175
		Tetrachloro-m-xylene	2	20	17.3	86		32	144
		Decachlorobiphenyl	2	20	16.4	82		32	175
P4255-01MS	OR-03-100124MS	Tetrachloro-m-xylene	1	20	17.1	85		32	144
		Decachlorobiphenyl	1	20	13.5	68		32	175
		Tetrachloro-m-xylene	2	20	14.9	75		32	144
		Decachlorobiphenyl	2	20	12.3	62		32	175
P4255-01MSD	OR-03-100124MSD	Tetrachloro-m-xylene	1	20	17.3	87		32	144
		Decachlorobiphenyl	1	20	13.1	65		32	175
		Tetrachloro-m-xylene	2	20	14.9	75		32	144
		Decachlorobiphenyl	2	20	12.3	61		32	175
I.BLK-PP067445.D	PIBLK-PP067445.D	Tetrachloro-m-xylene	1	20	16.5	82		60	140
		Decachlorobiphenyl	1	20	16.1	81		60	140
		Tetrachloro-m-xylene	2	20	16.4	82		60	140
		Decachlorobiphenyl	2	20	16.5	82		60	140
P4258-03	Chamberlain Ave	Tetrachloro-m-xylene	1	20	14.3	72		32	144
		Decachlorobiphenyl	1	20	13.1	66		32	175
		Tetrachloro-m-xylene	2	20	14.2	71		32	144
		Decachlorobiphenyl	2	20	13.5	68		32	175
I.BLK-PP067456.D	PIBLK-PP067456.D	Tetrachloro-m-xylene	1	20	16.4	82		60	140
		Decachlorobiphenyl	1	20	16.5	83		60	140
		Tetrachloro-m-xylene	2	20	16.3	81		60	140
		Decachlorobiphenyl	2	20	16.7	84		60	140