

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: ROMA02

Lab Code: CHEM **Case No.:** P4258 **SAS No.:** P4258 **SDG NO.:** P4258

Instrument ID: ECD_L **Date(s) Analyzed:** 09/23/2024 09/23/2024

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.24	6.14	6.34	22665300
		2	6.45	6.35	6.55	13922000
		3	7.06	6.96	7.16	70960500
		4	7.15	7.05	7.25	53851200
		5	7.94	7.84	8.04	39727000

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: ROMA02

Lab Code: CHEM **Case No.:** P4258 **SAS No.:** P4258 **SDG NO.:** P4258

Instrument ID: ECD_L **Date(s) Analyzed:** 09/23/2024 09/23/2024

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.01	4.91	5.11	18093600
		2	5.34	5.24	5.44	17271700
		3	6.61	6.51	6.71	63995900
		4	6.74	6.64	6.84	88666300
		5	7.05	6.95	7.15	60349700