

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: **ROMA02**

Lab Code: **CHEM** Case No.: **P4258** SAS No.: **P4258** SDG NO.: **P4258**

GC Column: **ZB-MR2** ID: **0.32** (mm) Initi. Calib. Date(s): **09/23/2024** **09/23/2024**

Client Sample No. (PEM): **PEM - PL091954.D** Date Analyzed: **09/23/2024**

Lab Sample No.(PEM): **PEM** Time Analyzed: **11:05**

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	9.061	8.960	9.160	19.830	20.000	-0.9
Tetrachloro-m-xylene	3.544	3.490	3.590	20.070	20.000	0.4
alpha-BHC	3.999	3.950	4.050	10.180	10.000	1.8
beta-BHC	4.529	4.480	4.580	10.750	10.000	7.5
gamma-BHC (Lindane)	4.332	4.280	4.380	10.100	10.000	1.0
Endrin	6.579	6.510	6.650	43.290	50.000	-13.4
4,4'-DDT	7.028	6.960	7.100	90.240	100.000	-9.8
Methoxychlor	7.504	7.430	7.570	212.390	250.000	-15.0

GC Column: **ZB-MR1** ID: **0.32** (mm) Initi. Calib. Date(s): **09/23/2024** **09/23/2024**

Client Sample No. (PEM): **PEM - PL091954.D** Date Analyzed: **09/23/2024**

Lab Sample No.(PEM): **PEM** Time Analyzed: **11:05**

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	7.923	7.820	8.020	19.350	20.000	-3.3
Tetrachloro-m-xylene	2.782	2.730	2.830	19.290	20.000	-3.6
alpha-BHC	3.285	3.230	3.340	9.220	10.000	-7.8
beta-BHC	3.915	3.860	3.970	10.540	10.000	5.4
gamma-BHC (Lindane)	3.616	3.570	3.670	8.880	10.000	-11.2
Endrin	5.648	5.580	5.720	46.450	50.000	-7.1
4,4'-DDT	6.047	5.980	6.120	104.950	100.000	5.0
Methoxychlor	6.621	6.550	6.690	232.550	250.000	-7.0