

Surrogate Summary

SDG No.: **P4258**

Client: **Roman E&G Corp**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL091953.D	PIBLK-PL091953.D	Decachlorobiphenyl	1	20	19.6	98		43	140
		Tetrachloro-m-xylene	1	20	19.0	95		77	126
		Decachlorobiphenyl	2	20	18.7	94		43	140
		Tetrachloro-m-xylene	2	20	18.3	92		77	126
I.BLK-PL092329.D	PIBLK-PL092329.D	Decachlorobiphenyl	1	20	23.8	119		43	140
		Tetrachloro-m-xylene	1	20	21.3	107		77	126
		Decachlorobiphenyl	2	20	22.1	110		43	140
		Tetrachloro-m-xylene	2	20	20.6	103		77	126
PB164067BL	PB164067BL	Decachlorobiphenyl	1	20	19.6	98		10	148
		Tetrachloro-m-xylene	1	20	19.0	95		10	159
		Decachlorobiphenyl	2	20	19.6	98		10	148
		Tetrachloro-m-xylene	2	20	18.1	90		10	159
PB164067BS	PB164067BS	Decachlorobiphenyl	1	20	22.0	110		10	148
		Tetrachloro-m-xylene	1	20	20.7	103		10	159
		Decachlorobiphenyl	2	20	21.3	106		10	148
		Tetrachloro-m-xylene	2	20	19.9	100		10	159
P4258-03	Chamberlain Ave	Decachlorobiphenyl	1	20	19.7	98		10	148
		Tetrachloro-m-xylene	1	20	19.9	99		10	159
		Decachlorobiphenyl	2	20	18.7	93		10	148
		Tetrachloro-m-xylene	2	20	19.4	97		10	159
P4385-04MS	SP-2MS	Decachlorobiphenyl	1	20	16.9	85		10	148
		Tetrachloro-m-xylene	1	20	16.4	82		10	159
		Decachlorobiphenyl	2	20	15.7	78		10	148
		Tetrachloro-m-xylene	2	20	16.3	81		10	159
P4385-04MSD	SP-2MSD	Decachlorobiphenyl	1	20	16.8	84		10	148
		Tetrachloro-m-xylene	1	20	16.6	83		10	159
		Decachlorobiphenyl	2	20	15.6	78		10	148
		Tetrachloro-m-xylene	2	20	15.9	79		10	159
I.BLK-PL092346.D	PIBLK-PL092346.D	Decachlorobiphenyl	1	20	21.1	105		43	140
		Tetrachloro-m-xylene	1	20	21.4	107		77	126
		Decachlorobiphenyl	2	20	19.7	99		43	140
		Tetrachloro-m-xylene	2	20	21.2	106		77	126