

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

Chamberlain AveMS

Contract: ROMA02

Lab Code: CHEM

Case No.: P4258

SAS No.: P4258

SDG NO.: P4258

Lab Sample ID: P4258-04MS

Date(s) Analyzed: 10/04/2024 10/04/2024

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP ID: 0.32 (mm)

GC Column:(2): RTX-CLP2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4-D	1	8.18	8.13	8.23	51.9	13.1
	2	8.85	8.80	8.90	59.2	
2,4,5-TP(Silvex)	1	9.04	8.99	9.09	49.0	27.5
	2	9.74	9.69	9.79	64.6	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

Chamberlain AveMSD

Contract: ROMA02

Lab Code: CHEM

Case No.: P4258

SAS No.: P4258

SDG NO.: P4258

Lab Sample ID: P4258-04MSD

Date(s) Analyzed: 10/04/2024

10/04/2024

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP

ID: 0.32 (mm)

GC Column:(2): RTX-CLP2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4-D	1	8.19	8.14	8.24	53.3	10.8
	2	8.85	8.80	8.90	59.4	
2,4,5-TP(Silvex)	1	9.04	8.99	9.09	50.4	38.5
	2	9.74	9.69	9.79	74.4	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB163890BS

Contract: ROMA02

Lab Code: CHEM

Case No.: P4258

SAS No.: P4258

SDG NO.: P4258

Lab Sample ID: PB163890BS

Date(s) Analyzed: 10/04/2024 10/04/2024

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP ID: 0.32 (mm)

GC Column:(2): RTX-CLP2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4-D	1	8.19	8.14	8.24	5.30	10.7
	2	8.85	8.80	8.90	5.90	
2,4,5-TP(Silvex)	1	9.04	8.99	9.09	5.20	3.9
	2	9.74	9.69	9.79	5.00	