

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

Instrument ID: ECD_S

Calibration Date(s): 09/12/2024 09/12/2024

Calibration Times: 21:24 22:59

GC Column: RTX-CLP ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS027653.D</u>	CF 500 =	<u>PS027654.D</u>		
CF 750 =		<u>PS027655.D</u>	CF 1000 =	<u>PS027656.D</u>	CF 1500 =	<u>PS027657.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	16480400000	15003800000	14421200000	14173600000	13096000000	14635000000	8
2,4-D	3492740000	3113810000	2979600000	2958000000	2782030000	3065240000	9
2,4-DCAA	2882370000	2518620000	2408720000	2392240000	2215400000	2483470000	10

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

Instrument ID: ECD_S Calibration Date(s): 09/12/2024 09/12/2024
Calibration Times: 21:24 22:59

GC Column: RTX-CLP2 ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS027653.D</u>	CF 500 =	<u>PS027654.D</u>		
CF 750 =		<u>PS027655.D</u>	CF 1000 =	<u>PS027656.D</u>	CF 1500 =	<u>PS027657.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	5828170000	5909010000	6118970000	6127280000	5996050000	5995900000	2
2,4-D	1161930000	1098640000	1103250000	1118490000	1092760000	1115020000	3
2,4-DCAA	1082400000	978389000	981157000	972356000	940861000	991033000	5