

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM **Case No.:** P4258 **SAS No.:** P4258 **SDG NO.:** P4258

Continuing Calib Date: 10/04/2024 **Initial Calibration Date(s):** 09/12/2024 09/12/2024

Continuing Calib Time: 04:09 **Initial Calibration Time(s):** 21:24 22:59

GC Column: RTX-CLP **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.10	7.11	7.01	7.21	0.01
2,4-D	8.19	8.20	8.10	8.30	0.01
2,4,5-TP(Silvex)	9.04	9.06	8.96	9.16	0.02

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM **Case No.:** P4258 **SAS No.:** P4258 **SDG NO.:** P4258

Continuing Calib Date: 10/04/2024 **Initial Calibration Date(s):** 09/12/2024 09/12/2024

Continuing Calib Time: 04:09 **Initial Calibration Time(s):** 21:24 22:59

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.62	7.63	7.53	7.73	0.01
2,4-D	8.85	8.86	8.76	8.96	0.01
2,4,5-TP(Silvex)	9.74	9.76	9.66	9.86	0.02