

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 09/12/2024 09/12/2024

Client Sample No.: CCAL01 Date Analyzed: 10/03/2024

Lab Sample No.: HSTDCCC750 Data File : PS027828.D Time Analyzed: 21:23

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.039	8.956	9.156	740.020	712.500	3.9
2,4-D	8.186	8.101	8.301	712.760	705.000	1.1
2,4-DCAA	7.099	7.012	7.212	748.300	750.000	-0.2

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/12/2024 09/12/2024

Client Sample No.: CCAL01 Date Analyzed: 10/03/2024

Lab Sample No.: HSTDCCC750 Data File : PS027828.D Time Analyzed: 21:23

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.737	9.655	9.855	729.370	712.500	2.4
2,4-D	8.848	8.761	8.961	714.120	705.000	1.3
2,4-DCAA	7.622	7.533	7.733	754.600	750.000	0.6