

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 09/12/2024 09/12/2024

Client Sample No.: CCAL02 Date Analyzed: 10/04/2024

Lab Sample No.: HSTDCCC750 Data File : PS027842.D Time Analyzed: 04:09

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.037	8.956	9.156	726.660	712.500	2.0
2,4-D	8.185	8.101	8.301	694.740	705.000	-1.5
2,4-DCAA	7.098	7.012	7.212	754.310	750.000	0.6

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/12/2024 09/12/2024

Client Sample No.: CCAL02 Date Analyzed: 10/04/2024

Lab Sample No.: HSTDCCC750 Data File : PS027842.D Time Analyzed: 04:09

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.736	9.655	9.855	745.500	712.500	4.6
2,4-D	8.847	8.761	8.961	718.340	705.000	1.9
2,4-DCAA	7.621	7.533	7.733	758.460	750.000	1.1