

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS091224

Review By	Abdul	Review On	9/15/2024 7:48:35 PM
Supervise By	Ankita	Supervise On	9/16/2024 10:13:53 AM
SubDirectory	PS091224	HP Acquire Method	HP Processing Method ps091224 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469		
CCC Internal Standard/PEM	PP23462		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP23469		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PS027651.D	12 Sep 2024 20:36	AR\AJ	Ok
2	I.BLK	PS027652.D	12 Sep 2024 21:00	AR\AJ	Ok
3	HSTDICC200	PS027653.D	12 Sep 2024 21:24	AR\AJ	Ok,M
4	HSTDICC500	PS027654.D	12 Sep 2024 21:48	AR\AJ	Ok
5	HSTDICC750	PS027655.D	12 Sep 2024 22:12	AR\AJ	Ok
6	HSTDICC1000	PS027656.D	12 Sep 2024 22:35	AR\AJ	Ok
7	HSTDICC1500	PS027657.D	12 Sep 2024 22:59	AR\AJ	Ok
8	HSTDICV750	PS027658.D	12 Sep 2024 23:23	AR\AJ	Ok,M
9	I.BLK	PS027659.D	12 Sep 2024 23:47	AR\AJ	Ok
10	HSTDCCC750	PS027660.D	13 Sep 2024 00:11	AR\AJ	Ok
11	P3845-17	PS027661.D	13 Sep 2024 00:35	AR\AJ	Ok
12	I.BLK	PS027662.D	13 Sep 2024 00:59	AR\AJ	Ok
13	HSTDCCC750	PS027663.D	13 Sep 2024 01:22	AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS100324

Review By	Abdul	Review On	10/4/2024 9:10:14 AM
Supervise By	Ankita	Supervise On	10/4/2024 9:35:20 AM
SubDirectory	PS100324	HP Acquire Method	HP Processing Method ps091224 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469		
CCC Internal Standard/PEM	PP23462		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP23469		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PS027814.D	03 Oct 2024 11:13	AR\AJ	Ok
2	I.BLK	PS027815.D	03 Oct 2024 11:37	AR\AJ	Ok,M
3	HSTDCCC750	PS027816.D	03 Oct 2024 12:01	AR\AJ	Ok
4	PB163866BL	PS027817.D	03 Oct 2024 17:01	AR\AJ	Ok
5	PB163866BS	PS027818.D	03 Oct 2024 17:25	AR\AJ	Ok,M
6	P4222-01	PS027819.D	03 Oct 2024 17:48	AR\AJ	Ok,M
7	P4222-01MS	PS027820.D	03 Oct 2024 18:12	AR\AJ	Ok,M
8	P4222-01MSD	PS027821.D	03 Oct 2024 18:36	AR\AJ	Ok,M
9	P4232-01	PS027822.D	03 Oct 2024 19:00	AR\AJ	ReRun
10	P4236-01	PS027823.D	03 Oct 2024 19:24	AR\AJ	Ok,M
11	P4236-05	PS027824.D	03 Oct 2024 19:48	AR\AJ	Ok,M
12	P4236-09	PS027825.D	03 Oct 2024 20:12	AR\AJ	Ok,M
13	P4236-13	PS027826.D	03 Oct 2024 20:36	AR\AJ	Ok,M
14	I.BLK	PS027827.D	03 Oct 2024 20:59	AR\AJ	Ok
15	HSTDCCC750	PS027828.D	03 Oct 2024 21:23	AR\AJ	Ok,M
16	P4276-01	PS027829.D	03 Oct 2024 22:59	AR\AJ	Ok,M
17	P4277-01	PS027830.D	03 Oct 2024 23:23	AR\AJ	Ok,M
18	P4242-01	PS027831.D	03 Oct 2024 23:47	AR\AJ	Ok,M
19	P4254-01	PS027832.D	04 Oct 2024 00:11	AR\AJ	Ok,M
20	P4270-01	PS027833.D	04 Oct 2024 00:34	AR\AJ	Ok,M
21	P4273-01	PS027834.D	04 Oct 2024 00:58	AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QCBatch ID # PS100324

Review By	Abdul	Review On	10/4/2024 9:10:14 AM
Supervise By	Ankita	Supervise On	10/4/2024 9:35:20 AM
SubDirectory	PS100324	HP Acquire Method	HP Processing Method ps091224 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469		
CCC Internal Standard/PEM	PP23462		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP23469		

22	P4258-04	PS027835.D	04 Oct 2024 01:22	AR\AJ	Ok,M
23	P4258-04MS	PS027836.D	04 Oct 2024 01:46	AR\AJ	Ok,M
24	P4258-04MSD	PS027837.D	04 Oct 2024 02:10	AR\AJ	Ok,M
25	PB163890BL	PS027838.D	04 Oct 2024 02:34	AR\AJ	Ok
26	PB163890BS	PS027839.D	04 Oct 2024 02:58	AR\AJ	Ok,M
27	PB163851TB	PS027840.D	04 Oct 2024 03:22	AR\AJ	Ok,M
28	I.BLK	PS027841.D	04 Oct 2024 03:45	AR\AJ	Ok,M
29	HSTDCCC750	PS027842.D	04 Oct 2024 04:09	AR\AJ	Ok,M

M : Manual Integration