

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

Instrument ID: ECD_L Calibration Date(s): 09/23/2024 09/23/2024
Calibration Times: 11:32 12:26

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 100 = <u>PL091956.D</u>	CF 075 = <u>PL091957.D</u>				
CF 050 = <u>PL091958.D</u>		CF 025 = <u>PL091959.D</u>	CF 005 = <u>PL091960.D</u>				
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
Decachlorobiphenyl	1514180000	1546290000	1590790000	1642930000	1896960000	1638230000	9
Endrin	1940880000	1934170000	1971650000	2016300000	2440480000	2060700000	10
gamma-BHC (Lindane)	2990410000	2948600000	2949110000	2954630000	3651280000	3098800000	10
Heptachlor	2590450000	2581810000	2624270000	2731930000	3298830000	2765460000	11
Heptachlor epoxide	2355250000	2419410000	2410600000	2493740000	3045980000	2544990000	11
Methoxychlor	913738000	926950000	954169000	973623000	1140080000	981712000	9
Tetrachloro-m-xylene	2128490000	2125270000	2145440000	2197480000	2526630000	2224660000	8

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

Instrument ID: ECD_L Calibration Date(s): 09/23/2024 09/23/2024
Calibration Times: 11:32 12:26

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PL091956.D</u>	CF 075 =	<u>PL091957.D</u>		
CF 050 =		<u>PL091958.D</u>	CF 025 =	<u>PL091959.D</u>	CF 005 =	<u>PL091960.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
Decachlorobiphenyl	2374340000	2399600000	2428470000	2478660000	2828640000	2501940000	7
Endrin	2724440000	2642180000	2624570000	2477520000	2642190000	2622180000	3
gamma-BHC (Lindane)	3775710000	3629090000	3577820000	3312530000	3359720000	3530980000	5
Heptachlor	3553790000	3434350000	3431490000	3238430000	3595010000	3450610000	4
Heptachlor epoxide	3090900000	2996730000	3012410000	2873000000	3141390000	3022890000	3
Methoxychlor	1228030000	1222750000	1234080000	1223180000	1343690000	1250350000	4
Tetrachloro-m-xylene	2544530000	2497760000	2476670000	2384990000	2636430000	2508080000	4